



MS ISO/IEC 17025

Certificate of Accreditation

No: SMM 088

Valid until: 14 September 2014

This is to certify that

**SIME-SIRIM TECHNOLOGIES SDN. BHD.
SUBANG JAYA, SELANGOR
MALAYSIA**

(FIELDS OF CALIBRATION: ELECTRICAL, TEMPERATURE, MASS,
DIMENSIONAL, FORCE, PRESSURE, TORQUE, VOLUMETRIC &
HYDROMETER)

has been granted accreditation in respect of the scope of accreditation described in the SCHEDULE attached, subject to the terms and conditions governing the *Skim Akreditasi Makmal Malaysia (SMM)*, the Laboratory Accreditation Scheme of Malaysia.

Laboratories accredited under SMM meet the requirements of MS ISO/IEC 17025 'General requirements for the competence of testing and calibration laboratories'. This Malaysian Standard is identical with ISO/IEC 17025 published by the International Organization for Standardization (ISO).

"This laboratory is accredited in accordance with recognized International Standard ISO/IEC 17025:2005. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer joint ISO-ILAC-IAF Communique dated 18 June 2005)"



(FADILAH BAHARIN)

Director General

Department of Standards Malaysia

Date of issue: 12 September 2011

NO : SAMM 088

Page: 1 of 60

LABORATORY LOCATION:
(PERMANENT LABORATORY)SIME-SIRIM TECHNOLOGIES SDN. BHD.
KOMPLEKS SIME DARBY
PERSIARAN KEWAJIPAN
47600 SUBANG JAYA, SELANGOR
MALAYSIA

This laboratory accredited under *Skim Akreditasi Makmal Malaysia* (SAMM) meets the requirements of MS ISO/IEC 17025:2005 'General requirements for competence of testing and calibration laboratories'. This Malaysian Standards is identical with ISO/IEC 17025:2005 published by the International Organization for Standardization (ISO).

* The expanded uncertainties are based on an estimated confidence probability of not less than 95% and have a coverage factor of $k=2$ unless stated otherwise.

FIELD OF CALIBRATION: ELECTRICAL

SCOPE OF ACCREDITATION:

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
1. Measuring Instrument DC VOLTAGE	± 220 mV Range \pm (0 mV to 220 mV)	9 μ V/V + 0.8 μ V	Generating using calibrator model Fluke 5700A
	± 2.2 V Range \pm (0 V to 2.2 V)	8 μ V/V + 1.2 μ V	
	+11 V Range \pm (0 V to 11 V)	8 μ V/V + 4 μ V	
	± 22 V Range \pm (0 V to 22 V)	8 μ V/V + 8 μ V	
	± 220 V Range \pm (0 V to 220 V)	9 μ V/V + 0.1 mV	
	± 1100 V Range \pm (100 V to 1100 V)	11 μ V/V + 0.6 mV	

NO : SAMM 088

Page: 2 of 60

FIELD OF CALIBRATION: ELECTRICAL

SCOPE OF ACCREDITATION:

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
1. Measuring Instrument AC VOLTAGE	<u>2.2 mV to 220 V</u> (See Matrix A)	(See Matrix A)	Generating using calibrator model Fluke 5700A
	<u>1100V Range (110V to 1100 V)</u> 50 Hz to 1 kHz	90 μ V/V + 4 mV	
	<u>800.01 V to 1050.00 V</u> 1 kHz to 3 kHz 3 kHz to 10 kHz 10 kHz to 20 kHz	0.8 mV/V + 0.13 V 0.8 mV/V + 0.21 V 1.2 mV/V + 0.32 V	Generating using calibrator model Wavetek 9100
DC CURRENT	0 mA to +2.2 mA -2.2 mA to 0 mA	60 μ A/A + 0.01 μ A	Generating using calibrator model Fluke 5700A
	0 mA to +22 mA -22 mA to 0 mA	60 μ A/A + 0.1 μ A	
	0 mA to +220 mA -220 mA to 0 mA	70 μ A/A + 1 μ A	
	0 A to +1A -1 A to 0 A	95 μ A/A + 30 μ A	Generating using calibrator model Wavetek 9100
	+0.32001 A to +3.2 A -3.2 A to -0.32001 A	0.6 mA/A + 0.12 mA	
	+3.2001 A to +10.5 A -10.5 A to -3.2001 A	0.55 mA/A + 0.94 mA	
	+10.5001 A to +20 A -20 A to -10.5001 A	0.55 mA/A + 4.5 mA	
	+20 A to +100 A -100 A to -20 A	0.45 mA/A + 20 mA	California 3213K



NO : SAMM 088

Page: 3 of 60

FIELD OF CALIBRATION: ELECTRICAL

SCOPE OF ACCREDITATION:

Matrix A
AC Voltage Measurement

Range		Frequency			
		Hz		kHz	
2.2 mV	0.22 mV to 2.2 mV	0.6 + 0.005	0.24 + 0.005	0.12 + 0.005	0.41 + 0.005
22 mV	2.2 mV to 22 mV	0.6 + 0.006	0.24 + 0.006	0.12 + 0.006	0.41 + 0.006
220 mV	22 mV to 220 mV	0.6 + 0.016	0.24 + 0.01	0.11 + 0.01	0.36 + 0.01
2.2 V	0.22 V to 2.2 V	0.6 + 0.1	0.18 + 0.03	0.085 + 0.007	0.14 + 0.02
22 V	2.2 V to 22 V	0.6 + 1	0.18 + 0.3	0.085 + 0.07	0.14 + 0.2
220 V	22 V to 220 V	0.6 + 10	0.18 + 3	0.09 + 1	0.25 + 4

Range		Frequency			
		MHz			
2.2 mV	0.22 mV to 2.2 mV	0.95 + 0.008	1.3 + 0.015	1.8 + 0.03	3.6 + 0.03
22 mV	2.2 mV to 22 mV	0.95 + 0.008	1.3 + 0.015	1.8 + 0.03	3.6 + 0.03
220 mV	22 mV to 220 mV	0.9 + 0.03	1.1 + 0.03	1.8 + 0.04	3.6 + 0.1
2.2 V	0.22 V to 2.2 V	0.28 + 0.08	0.48 + 0.15	1.2 + 0.4	2.4 + 1
22 V	2.2 V to 22 V	0.28 + 0.4	0.6 + 1.7	1.4 + 5	3 + 9
220 V	22 V to 220 V	0.6 + 10	1.6 + 110	5.4 + 110	13 + 220

The expanded uncertainties given in this table are expressed in mV/V + mV



NO : SAMM 088

Page: 4 of 60

FIELD OF CALIBRATION: ELECTRICAL

SCOPE OF ACCREDITATION:

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
1. Measuring Instrument AC CURRENT	<u>9 μA to 22 A</u> (See Matrix B)	(See Matrix B)	Generating using calibrator model Fluke 5700A
	<u>0.32001 A to 3.2 A</u> 10 Hz to 3 kHz 3 kHz to 10 kHz	1 mA/A + 0.48 mA 2.5 mA/A + 2.6 mA	Generating using calibrator model Wavetek 9100
	<u>3.2001 A to 10.5 A</u> 10 Hz to 3 kHz 3 kHz to 10 kHz	2 mA/A + 3 mA 5 mA/A + 10 mA	
	<u>10.5001 A to 20 A</u> 10 Hz to 3 kHz 3 kHz to 10 kHz	2 mA/A + 6.9 mA 5 mA/A + 23 mA	
	<u>20 A to 100 A</u> 10 Hz to 1 kHz 1 kHz to 10 kHz	2.5 mA/A + 0.03 A 5 mA/A + 0.05 A	CA 3213K

Matrix B
AC Current Measurement

Range	Frequency (kHz)				
	0.01 to 0.02	0.02 to 0.04	0.04 to 1	1 to 5	5 to 10
9 μ A to 220 μ A	0.8 + 0.03	0.42 + 0.025	0.16 + 0.02	0.7 + 0.05	1.8 + 0.1
0.22 mA to 2.2 mA	0.8 + 0.05	0.42 + 0.04	0.16 + 0.04	0.7 + 0.5	1.8 + 1
2.2 mA to 22 mA	0.8 + 0.5	0.42 + 0.4	0.16 + 0.4	0.7 + 10	1.8 + 0.01
22 mA to 220 mA	0.8 + 5	0.42 + 4	0.18 + 4	0.7 + 50	1.8 + 100
0.22 A to 2.2 A	-	-	0.75 + 40	0.85 + 100	10 + 200

The expanded uncertainties given in this table are expressed in mA/A + μ A

NO : SAMM 088

Page: 5 of 60

FIELD OF CALIBRATION: ELECTRICAL

SCOPE OF ACCREDITATION:

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
RESISTANCE Fixed Value	1 m Ω 10 m Ω 100 m Ω 1 Ω 1.9 Ω 10 Ω 19 Ω 100 Ω 190 Ω 1 k Ω 1.9 k Ω 10 k Ω 19 k Ω 100 k Ω 190 k Ω 1 M Ω 1.9 M Ω 10 M Ω 19 M Ω	0.2 m Ω/Ω 0.1 m Ω/Ω 0.02 m Ω/Ω 5 $\mu\Omega/\Omega$ 0.11 m Ω/Ω 5 $\mu\Omega/\Omega$ 31 $\mu\Omega/\Omega$ 5 $\mu\Omega/\Omega$ 0.02 m Ω/Ω 5 $\mu\Omega/\Omega$ 15 $\mu\Omega/\Omega$ 5 $\mu\Omega/\Omega$ 14 $\mu\Omega/\Omega$ 5 $\mu\Omega/\Omega$ 15 $\mu\Omega/\Omega$ 5 $\mu\Omega/\Omega$ 24 $\mu\Omega/\Omega$ 5 $\mu\Omega/\Omega$ 0.05 m Ω/Ω	Fluke 5700A & Fluke 742A Series
	100 M Ω 1 G Ω 10 G Ω 100 G Ω 1 T Ω	0.13 m Ω/Ω 5 $\mu\Omega/\Omega$ 0.02 Ω/Ω 0.05 Ω/Ω 0.05 Ω/Ω	Fluke 5700A & Takeda Riken TR45
CAPACITANCE Fixed Value	<u>20 Hz to 1 MHz</u> 1 pF, 10 pF, 100 pF, 1000 pF, 0.01 μ F, 0.1 μ F and 1.0 μ F	1 mF/F	HP 1600A series
	<u>≤ 350 Hz</u> 0.5000 nF to 4.0000 nF 4.0001 nF to 40.000 nF 40.001 nF to 400.00 nF 400.01 nF to 4.0000 μ F 4.0001 μ F to 40.000 μ F 40.001 μ F to 400.00 μ F 400.01 μ F to 4.0000 mF 4.0001 mF to 40.000 mF	3 mF/F + 15 pF 3 mF/F + 30 pF 3 mF/F + 0.16 nF 4 mF/F + 1.6 nF 5 mF/F + 16 nF 5 mF/F + 0.16 μ F 5 mF/F + 1.6 μ F 10 mF/F + 0.06 mF	Generating using calibrator model Wavetek 9100

NO : SAMM 088

Page: 6 of 60

FIELD OF CALIBRATION: ELECTRICAL

SCOPE OF ACCREDITATION:

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
1. Measuring Instrument CAPACITANCE Fixed Value (continued)	350 Hz to 1.5 kHz 0.5000 nF to 4.0000 nF 4.0001 nF to 40.000 nF 40.001 nF to 400.00 nF 400.01 nF to 4.0000 μ F 4.0001 μ F to 40.000 μ F 40.001 μ F to 400.00 μ F 400.01 μ F to 4.0000 mF 4.0001 mF to 40.000 mF	6 mF/F + 0.03 nF 6 mF/F + 0.06 nF 6 mF/F + 0.32 nF 8 mF/F + 3.2 nF 10 mF/F + 32 nF 10 mF/F + 0.32 μ F 10 mF/F + 3.2 μ F 20 mF/F + 0.12 mF	Generating using calibrator model Wavetek 9100
INDUCTANCE Fixed Value	100 μ H @ 1 kHz (1, 10, 100) mH @ 1 kHz 1 H @ 1 kHz	2.5 mH/H 1 mH/H 1 mH/H	Ando RS-100 series
POWER/ ENERGY (DC)	1 kW to 20 kW 0.1 W to 1 kW	0.7 mW/W 0.22 mW/W	Generating using calibrator model Fluke 5520A.
POWER/ ENERGY(AC) (45 Hz to 65 Hz at PF=1)	10 kW to 20 kW 1 W to 10 kW 0.1 W to 1 W	1 mW/W 0.9 mW/W 1 mW/W	
FREQUENCY	1 μ Hz to 80 MHz 100 kHz to 4320 MHz	64 nHz/Hz 32 nHz/Hz	(Agilent 33250A & R&S SMHU) & HP 53132A
Time Base Output	1 kHz to 10 MHz	3.6 nHz/Hz	HP 53132A
AMPLITUDE	10 mVp-p to 10 Vp-p (1 μ Hz to 80 MHz, Into 50 Ω)	10 mVpp/Vpp + 1 mVpp	Agilent 33250A
	+13 dBm to -140 dBm (100 kHz to 4320 MHz)	1 dBm	R&S SMHU
FLATNESS	Sine Wave Relative to 1 kHz (Auto range On)	0.4 dBm	Agilent 33250A
TIME	1 s to 24 Hrs	0.06 μ s/s + 0.021 s	Agilent 33250A & HP 53132A



NO : SAMM 088

Page: 7 of 60

FIELD OF CALIBRATION: ELECTRICAL

SCOPE OF ACCREDITATION:

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
1. Measuring Instrument RPM Measuring Instruments (Non-Contact)	60 to 5999.9 rpm 6000 to 29999 rpm 30000 to 59999 rpm 60000 to 99999 rpm	0.07 rpm 7 rpm 36 rpm 70 rpm	In-House method, ESF/0303.
2. High Voltage Meter DC VOLTAGE	0.5 kV to 10 kV	5 mV/V	TDV 20 ADS & TOS 5101
AC VOLTAGE	(0.5 kV to 10 kV) @ (50/60) Hz	8 mV/V	
3. Clamp Meter DC Current <u>10 – Turn Coil</u>	3.2001 A to 32 A 32.001 A to 105 A 105.001 A to 200 A	0.06 mA/A + 1.18 mA 0.55 mA/A + 9.4 mA 0.55 mA/A + 45 mA	Generating using calibrator model Wavetek 9100 c/w Current Coil
<u>50 – Turn Coil</u>	16.001 A to 160 A 160.01 A to 525.0 A 525.01 A to 1000.0 A	0.6 mA/A + 5.9 mA 0.55 mA/A + 47 mA 0.55 mA/A + 0.23 A	
AC Current <u>10 – Turn Coil</u>	<u>3.2001 A to 32.0 A</u> 10 Hz to 100 Hz 100 Hz to 440 Hz <u>32.001 A to 200.0 A</u> 10 Hz to 100 Hz 100 Hz to 440 Hz	 2 mA/A + 5.5 mA 7.8 mA/A + 27 mA 2.1 mA/A + 90 mA 6.7 mA/A + 0.25 A	
<u>50 – Turn Coil</u>	<u>16.001 A to 160.0 A</u> 10 Hz to 100 Hz <u>160.01 A to 1000.0 A</u> 10 Hz to 100 Hz	 2 mA/A + 28 mA 2.1 mA/A + 0.45 A	



NO : SAMM 088

Page: 8 of 60

FIELD OF CALIBRATION: ELECTRICAL

SCOPE OF ACCREDITATION:

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
4. Insulation Testers	(1.0k Ω to 10.0 k Ω) @ 10.0 V (10.0k Ω to 100.0 k Ω) @ 50.0V (0.1M Ω to 1.0 M Ω) @ 150.0 V (1.0M Ω to 10.0M Ω) @ 300.0V (10.0M Ω to 100.0M Ω)@500.0V (0.1G Ω to 1.0 G Ω) @ 1000.0 V (1.0G Ω to 10.0G Ω) @ 5000.0V (10.0G Ω to 100.0G Ω) @ 5000.0V (100.0G Ω to 600.0G Ω) @ 5000.0V	0.5 m Ω / Ω 1 m Ω / Ω 10 m Ω / Ω 50 m Ω / Ω 0.1 Ω / Ω 50 m Ω / Ω	Tinsley 4720
5. Oscilloscope Vertical Deflection DC Signal	0V to \pm 6.6 V (50 Ω Load) 0V to \pm 130 V (1 M Ω Load)	2.5 mV/V + 0.04 mV 0.5 mV/V + 0.04 mV	Generating using calibrator model Fluke 5500A SC600
Vertical Deflection Square Wave Signal	\pm 1 mVp-p to \pm 6.6 Vp-p (50 Ω Load) \pm 1 mVp-p to \pm 130 Vp-p (1 M Ω Load)	2.5 mVp-p/Vp-p + 0.04 mVp-p 1 mVp-p/Vp-p + 0.04 mVp-p	
Horizontal Deflection Time Markers (50 Ω Load)	2 ns/div to 20 ms/div 50 ms/div to 5 s/div	2.5 μ s/s (25+(Output x 1000)) μ s/s	
Risetime	\leq 300 ps	+0 ns / -0.1 ns	
Bandwidth Frequency	50 kHz to 600 MHz	2.5 μ Hz/Hz	
	600 MHz to 4320 MHz	32 nHz/Hz	R&H SMHU &HP53132A
Bandwidth Amplitude	50 kHz to 600 MHz 600 MHz to 4320 MHz	0.03 Vp-p 0.26 Vp-p	Generating using calibrator model Fluke 5500A SC600 & R&H SMHU



NO : SAMM 088

Page: 9 of 60

FIELD OF CALIBRATION: ELECTRICAL

SCOPE OF ACCREDITATION:

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
6. Sourcing / Generating Instrument DC VOLTAGE	100 mV @ ± 100 % Band Span \pm (85 mV to 115 mV)	4 μ V/V	Measuring using Standard model Wavetek MTS 4950
	1 V @ ± 100 % Band Span \pm (0.9 V to 1.1 V)	2.2 μ V/V	
	10 V @ ± 100 % Band Span \pm (9 V to 11 V)	1.4 μ V/V	
	10 V @ ± 190 % Band Span \pm (18 V to 19.5 V)	1.8 μ V/V	
	100 V @ ± 100 % Band Span \pm (90 V to 110 V)	2 μ V/V	
	1000 V @ ± 100 % Band Span \pm (900 V to 1100 V)	2 μ V/V	
	± 100 mV Range \pm (100 μ V to 120 mV)	5 μ V/V + 0.3 μ V	Measuring using Multimeter model HP 3458A Opt 002
	± 1 V Range \pm (100 mV to 1.2 V)	4 μ V/V + 0.3 μ V	
	± 10 V Range \pm (1 V to 12 V)	4 μ V/V + 0.5 μ V	
	± 100 V Range \pm (10 V to 120 V)	6 μ V/V + 0.03 mV	
	± 1000 V Range \pm (100 V to 1050 V)	6 μ V/V + 0.1 mV	



NO : SAMM 088

Page: 10 of 60

FIELD OF CALIBRATION: ELECTRICAL

SCOPE OF ACCREDITATION:

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
6. Sourcing / Generating Instrument AC VOLTAGE	<u>1 mV @ 100 % Band Span</u> <u>(0.85 mV to 1.15 mV)</u>		
	10 Hz Frequency Band Span (9 Hz to 11 Hz),	0.28 mV/V	
	20 Hz Frequency Band Span (18 Hz to 22 Hz),		
	30 Hz Frequency Band Span (27 Hz to 33 Hz),		
	40 Hz Frequency Band Span (36 Hz to 44 Hz),		
	55 Hz Frequency Band Span (46.25 Hz to 63.75 Hz),		
	10 kHz Frequency Band Span (9 kHz to 11 kHz)		
	300 Hz Frequency Band Span (270 Hz to 440 Hz),	0.26 mV/V	
	1 kHz Frequency Band Span (0.9 kHz to 1.1 kHz)		
	20 kHz Frequency Band Span (18 kHz to 22 kHz)	0.3 mV/V	
	30 kHz Frequency Band Span (27 kHz to 33 kHz),	0.39 mV/V	
	50 kHz Frequency Band Span (45 kHz to 55 kHz)		
	100 kHz Frequency Band Span (90 kHz to 110 kHz)	0.62 mV/V	
	300 kHz Frequency Band Span (270 kHz to 330 kHz)	0.75 mV/V	
	500 kHz Frequency Band Span (450 kHz to 550 kHz)	0.77 mV/V	
	1 MHz Frequency Band Span (0.9 MHz to 1.1 MHz)	1.06 mV/V	
			Wavetek MTS 4950



NO : SAMM 088

Page: 11 of 60

FIELD OF CALIBRATION: ELECTRICAL

SCOPE OF ACCREDITATION:

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
6. Sourcing/ Generating instrument AC VOLTAGE (continue)	<u>2.2 mV Range</u> <u>(600 μV to 2.2 mV)</u> 10 Hz to 20 Hz 20 Hz to 40 Hz 40 Hz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 500 kHz 500 kHz to 1 MHz <u>7 mV Range (1.9 mV to 7 mV)</u> 10 Hz to 20 Hz 20 Hz to 40 Hz 40 Hz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 500 kHz 500 kHz to 1 MHz	1.7 mV/V + 1.3 μ V 0.74 mV/V + 1.3 μ V 0.42 mV/V + 1.3 μ V 0.81 mV/V + 2 μ V 1.2 mV/V + 2.5 μ V 2.3 mV/V + 4 μ V 2.4 mV/V + 8 μ V 3.5 mV/V + 8 μ V	Fluke 5790A
	<u>10 mV @ 100 % Band Span</u> <u>(8.5 mV to 11.5 mV)</u> 10 Hz Frequency Band Span (9 Hz to 11 Hz), 20 Hz Frequency Band Span (18 Hz to 22 Hz), 30 Hz Frequency Band Span (27 Hz to 33 Hz), 40 Hz Frequency Band Span (36 Hz to 44 Hz), 55 Hz Frequency Band Span (46.25 Hz to 63.75 Hz)	0.16 mV/V	Wavetek MTS 4950
	300 Hz Frequency Band Span (270 Hz to 440 Hz) & 1 kHz Frequency Band Span (0.9 kHz to 1.1 kHz)	0.14 mV/V	



NO : SAMM 088

Page: 12 of 60

FIELD OF CALIBRATION: ELECTRICAL

SCOPE OF ACCREDITATION:

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
6. Sourcing/ Generating Instrument AC VOLTAGE (continue)	<u>10 mV @ 100 % Band Span</u> <u>(8.5 mV to 11.5 mV)</u>		
	10 kHz Frequency Band Span (9 kHz to 11 kHz) & 20 kHz Frequency Band Span (18 kHz to 22 kHz)	0.16 mV/V	
	30 kHz Frequency Band Span (27 kHz to 33 kHz)	0.24 mV/V	
	50 kHz Frequency Band Span (45 kHz to 55 kHz), 100 kHz Frequency Band Span (90 kHz to 110 kHz), 300 kHz Frequency Band Span (270 kHz to 330 kHz)	0.59 mV/V	Wavetek MTS 4950
	500 kHz Frequency Band Span (450 kHz to 550 kHz)	0.61 mV/V	
	1 MHz Frequency Band Span (0.9 MHz to 1.1 MHz)	0.95 mV/V	
	<u>10 mV Range (1 mV to 12 mV)</u> 1 Hz to 40 Hz 1 MHz to 4 MHz 4 MHz to 8 MHz	0.3 mV/V + 3 μ V 0.07 V/V + 7 μ V 0.2 V/V + 8 μ V	HP 3458A Opt 002
	<u>22 mV Range (6 mV to 22 mV)</u> 10 Hz to 20 Hz 20 Hz to 40 Hz 40 Hz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 500 kHz 500 kHz to 1 MHz	0.29 mV/V + 1.3 μ V 0.19 mV/V + 1.3 μ V 0.11 mV/V + 1.3 μ V 0.21 mV/V + 2 μ V 0.31 mV/V + 2.5 μ V 0.81 mV/V + 4 μ V 0.89 mV/V + 8 μ V 1.7 mV/V + 8 μ V	Fluke 5790A

NO : SAMM 088

Page: 13 of 60

FIELD OF CALIBRATION: ELECTRICAL

SCOPE OF ACCREDITATION:

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
6. Sourcing/ Generating Instrument AC VOLTAGE (continue)	<u>70 mV Range (19 mV to 70 mV)</u> 10 Hz to 20 Hz 20 Hz to 40 Hz 40 Hz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 500 kHz 500 kHz to 1 MHz	0.24 mV/V + 1.5 μ V 0.12 mV/V + 1.5 μ V 65 μ V/V + 1.5 μ V 0.13 mV/V + 2 μ V 0.26 mV/V + 2.5 μ V 0.51 mV/V + 4 μ V 0.67 mV/V + 8 μ V 1.1 mV/V + 8 μ V	Fluke 5790A
	<u>100 mV @ 100 % Band Span</u> <u>(85 mV to 115 mV)</u> 10 Hz Frequency Band Span (9 Hz to 11 Hz), 20 Hz Frequency Band Span (18 Hz to 22 Hz), 30 Hz Frequency Band Span (27 Hz to 33 Hz), 40 Hz Frequency Band Span (36 Hz to 44 Hz) & 55 Hz Frequency Band Span (46.25 Hz to 63.75 Hz) 300 Hz Frequency Band Span (270 Hz to 440 Hz) & 1 kHz Frequency Band Span (0.9 kHz to 1.1 kHz) 10 kHz Frequency Band Span (9 kHz to 11 kHz) 20 kHz Frequency Band Span (18 kHz to 22 kHz) 30 kHz Frequency Band Span (27 kHz to 33 kHz) & 50 kHz Frequency Band Span (45 kHz to 55 kHz)	0.12 mV/V 89 μ V/V 0.11 mV/V 0.12 mV/V 0.19 mV/V	Wavetek MTS 4950



NO : SAMM 088

Page: 14 of 60

FIELD OF CALIBRATION: ELECTRICAL

SCOPE OF ACCREDITATION:

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
6. Sourcing/ Generating Instrument AC VOLTAGE (continue)	<u>100 mV @ 100 % Band Span</u> (85 mV to 115 mV)		
	100 kHz Frequency Band Span (90 kHz to 110 kHz)	0.36 mV/V	Wavetek MTS 4950
	300 kHz Frequency Band Span (270 kHz to 330 kHz)	96 μ V/V	
	500 kHz Frequency Band Span (450 kHz to 550 kHz)	0.61 mV/V	
	1 MHz Frequency Band Span (0.9 MHz to 1.1 MHz)	0.95 mV/V	
	<u>100 mV range (10 mV to 120 mV)</u>		
	1 Hz to 40 Hz	0.07 mV/V + 4 μ V	HP 3458A Opt 002
	1 MHz to 2 MHz	15 mV/V + 0.01 mV	
	2 MHz to 4 MHz	0.04 V/V + 0.07 mV	
	4 MHz to 8 MHz	0.04 V/V + 0.08 mV	
	8 MHz to 10 MHz	0.15 V/V + 0.1 mV	
	<u>220 mV Range (60 mV to 220 mV)</u>		
	10 Hz to 20 Hz	0.21 mV/V + 1.5 μ V	Fluke 5790A
	20 Hz to 40 Hz	85 μ V/V + 1.5 μ V	
	40 Hz to 20 kHz	38 μ V/V + 1.5 μ V	
	20 kHz to 50 kHz	69 μ V/V + 2 μ V	
	50 kHz to 100 kHz	0.16 mV/V + 2.5 μ V	
	100 kHz to 300 kHz	0.25 mV/V + 4 μ V	
	300 kHz to 500 kHz	0.38 mV/V + 8 μ V	
	500 kHz to 1 MHz	1 mV/V + 8 μ V	
	<u>700 mV Range</u> (190 mV to 700 mV)		
	10 Hz to 20 Hz	0.21 mV/V + 1.5 μ V	
	20 Hz to 40 Hz	76 μ V/V + 1.5 μ V	
	40 Hz to 20 kHz	33 μ V/V + 1.5 μ V	
	20 kHz to 50 kHz	51 μ V/V + 2 μ V	
	50 kHz to 100 kHz	79 μ V/V + 2.5 μ V	
	100 kHz to 300 kHz	0.18 mV/V + 4 μ V	
	300 kHz to 500 kHz	0.3 mV/V + 8 μ V	
	500 kHz to 1 MHz	0.96 mV/V + 8 μ V	



NO : SAMM 088

Page: 15 of 60

FIELD OF CALIBRATION: ELECTRICAL

SCOPE OF ACCREDITATION:

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
6. Sourcing/ Generating Instrument AC VOLTAGE (continue)	<u>1 V @ 100 % Band Span</u> (0.9 V to 1.1 V)		
	10 Hz Frequency Band Span (9 Hz to 11 Hz),	36 μ V/V	
	20 Hz Frequency Band Span (18 Hz to 22 Hz),		
	30 Hz Frequency Band Span (27 Hz to 33 Hz),		
	40 Hz Frequency Band Span (36 Hz to 44 Hz),	24 μ V/V	
	55 Hz Frequency Band Span (46.25 Hz to 63.75 Hz),		
	300 Hz Frequency Band Span (270 Hz to 440 Hz),		
	1 kHz Frequency Band Span (0.9 kHz to 1.1 kHz),		
	10 kHz Frequency Band Span (9 kHz to 11 kHz),		
	20 kHz Frequency Band Span (18 kHz to 22 kHz)		Wavetek MTS 4950
	30 kHz Frequency Band Span (27 kHz to 33 kHz)	24 μ V/V	
	50 kHz Frequency Band Span (45 kHz to 55 kHz)	31 μ V/V	
	100 kHz Frequency Band Span (90 kHz to 110 kHz)	37 μ V/V	
	300 kHz Frequency Band Span (270 kHz to 330 kHz)	96 μ V/V	
	500 kHz Frequency Band Span (450 kHz to 550 kHz)	0.21 mV/V	
	1 MHz Frequency Band Span (0.9 MHz to 1.1 MHz)	0.56 mV/V	

NO : SAMM 088

Page: 16 of 60

FIELD OF CALIBRATION: ELECTRICAL

SCOPE OF ACCREDITATION:

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
6. Sourcing/ Generating Instrument AC VOLTAGE (continue)	<u>1 V Range (0.1 V to 1.2 V)</u>		
	1 Hz to 40 Hz	0.07 mV/V + 0.04 mV	HP 3458A Opt 002
	1 MHz to 2 MHz	15 mV/V + 0.1 mV	
	2 MHz to 4 MHz	0.04 V/V + 0.7 mV	
	4 MHz to 8 MHz	0.04 V/V + 0.8 mV	
	8 MHz to 10 MHz	0.15 V/V + 1 mV	
	<u>2.2 V Range(600 mV to 2.2 V)</u>		
	10 Hz to 20 Hz	0.2 mV/V	Fluke 5790A
	20 Hz to 40 Hz	66 μ V/V	
	40 Hz to 20 kHz	24 μ V/V	
	20 kHz to 50 kHz	46 μ V/V	
	50 kHz to 100 kHz	71 μ V/V	
	100 kHz to 300 kHz	0.16 mV/V	
	300 kHz to 500 kHz	0.26 mV/V	
	500 kHz to 1 MHz	0.9 mV/V	
	<u>7 V Range (1.9 V to 7 V)</u>		
	10 Hz to 20 Hz	0.2 mV/V	
	20 Hz to 40 Hz	67 μ V/V	
	40 Hz to 20 kHz	24 μ V/V	
	20 kHz to 50 kHz	48 μ V/V	
	50 kHz to 100 kHz	81 μ V/V	
	100 kHz to 300 kHz	0.19 mV/V	
	300 kHz to 500 kHz	0.4 mV/V	
	500 kHz to 1 MHz	1.2 mV/V	
	<u>10 V @ 100 % Band Span (9 V to 11 V)</u>		
	10 Hz Frequency Band Span (9 Hz to 11 Hz),	36 μ V/V	Wavetek MTS 4950
	20 Hz Frequency Band Span (18 Hz to 22 Hz),		
	30 Hz Frequency Band Span (27 Hz to 33 Hz)		



NO : SAMM 088

Page: 17 of 60

FIELD OF CALIBRATION: ELECTRICAL

SCOPE OF ACCREDITATION:

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm) [*]	Remarks
6. Sourcing/ Generating Instrument AC VOLTAGE (continue)	<u>10 V @ 100 % Band Span</u> (9 V to 11 V)		
	40 Hz Frequency Band Span (36 Hz to 44 Hz),	24 μ V/V	
	55 Hz Frequency Band Span (46.25 Hz to 63.75 Hz),		
	300 Hz Frequency Band Span (270 Hz to 440 Hz),		
	1 kHz Frequency Band Span (0.9 kHz to 1.1 kHz),		
	10 kHz Frequency Band Span (9 kHz to 11 kHz),		
	20 kHz Frequency Band Span (18 kHz to 22 kHz)		
	30 kHz Frequency Band Span (27 kHz to 33 kHz),	26 μ V/V	
	50 kHz Frequency Band Span (45 kHz to 55 kHz),		
	100 kHz Frequency Band Span (90 kHz to 110 kHz)		
	300 kHz Frequency Band Span (270 kHz to 330 kHz)	83 μ V/V	
	500 kHz Frequency Band Span (450 kHz to 550 kHz)	0.19 mV/V	
	1 MHz Frequency Band Span (0.9 MHz to 1.1 MHz)	0.53 mV/V	
	<u>10 V @ 190 % Band Span</u> (18 V to 19.5 V)		
	1 kHz Frequency Band Span (0.9 kHz to 1.1 kHz)	24 μ V/V	
			Wavetek MTS 4950



NO : SAMM 088

Page: 18 of 60

FIELD OF CALIBRATION: ELECTRICAL

SCOPE OF ACCREDITATION:

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
6. Sourcing/ Generating Instrument AC VOLTAGE (continue)	<u>10 V Range (1 V to 12 V)</u>		
	1 Hz to 40 Hz	0.07 mV/V + 0.4 mV	HP 3458A Opt 002
	1 MHz to 2 MHz	15 mV/V + 1 mV	
	2 MHz to 4 MHz	0.04 V/V + 7 mV	
	4 MHz to 8 MHz	0.04 V/V + 8 mV	
	8 MHz to 10 MHz	0.15 V/V + 0.01 V	
	<u>22 V Range (6 V to 22 V)</u>		Fluke 5790A
	10 Hz to 20 Hz	0.2 mV/V	
	20 Hz to 40 Hz	67 μ V/V	
	40 Hz to 20 kHz	27 μ V/V	
	20 kHz to 50 kHz	48 μ V/V	
	50 kHz to 100 kHz	81 μ V/V	
	100 kHz to 300 kHz	0.19 mV/V	
	300 kHz to 500 kHz	0.4 mV/V	
	500 kHz to 1 MHz	1.2 mV/V	
	<u>70 V Range (19 V to 70 V)</u>		
	10 Hz to 20 Hz	0.2 mV/V	
	20 Hz to 40 Hz	68 μ V/V	
	40 Hz to 20 kHz	32 μ V/V	
	20 kHz to 50 kHz	57 μ V/V	
	50 kHz to 100 kHz	94 μ V/V	
	100 kHz to 300 kHz	0.2 mV/V	
	300 kHz to 500 kHz	0.41 mV/V	
	500 kHz to 1 MHz	1.2 mV/V	
	<u>100 V Range (10 V to 120 V)</u>		
	1 Hz to 40 Hz	0.2 mV/V + 4 mV	HP 3458A Opt 002
	<u>100 V @ 100 % Band Span (90 V to 110 V)</u>		Wavetek MTS 4950
	10 Hz Frequency Band Span (9 Hz to 11 Hz),	41 μ V/V	
	20 Hz Frequency Band Span (18 Hz to 22 Hz),		
	30 Hz Frequency Band Span (27 Hz to 33 Hz)		
	40 Hz Frequency Band Span (36 Hz to 44 Hz)	36 μ V/V	

NO : SAMM 088

Page: 19 of 60

FIELD OF CALIBRATION: ELECTRICAL

SCOPE OF ACCREDITATION:

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
6. Sourcing/ Generating Instrument AC VOLTAGE (continue)	<u>100 V @ 100 % Band Span</u> <u>(90 V to 110 V)</u>		
	55 Hz Frequency Band Span (46.25 Hz to 63.75 Hz)	36 μ V/V	
	300 Hz Frequency Band Span (270 Hz to 440 Hz), 1 kHz Frequency Band Span (0.9 kHz to 1.1 kHz), 10 kHz Frequency Band Span (9 kHz to 11 kHz), 20 kHz Frequency Band Span (18 kHz to 22 kHz)	26 μ V/V	Wavetek MTS 4950
	30 kHz Frequency Band Span (27 kHz to 33 kHz)	29 μ V/V	
	50 kHz Frequency Band Span (45 kHz to 55 kHz)	35 μ V/V	
	100 kHz Frequency Band Span (90 kHz to 110 kHz)	64 μ V/V	
	200 kHz Frequency Band Span (180 kHz to 220 kHz)	0.24 mV/V	
	<u>220 V Range (60 V to 220 V)</u>		Fluke 5790A
	10 Hz to 20 Hz	0.2 mV/V	
	20 Hz to 40 Hz	68 μ V/V	
	40 Hz to 20 kHz	31 μ V/V	
	20 kHz to 50 kHz	69 μ V/V	
	50 kHz to 100 kHz	98 μ V/V	
	100 kHz to 300 kHz	0.21 mV/V	
	300 kHz to 500 kHz	0.5 mV/V	
	<u>700 V Range (190 V to 700 V)</u>		
	10 Hz to 20 Hz	0.2 mV/V	
	20 Hz to 40 Hz	99 μ V/V	
	40 Hz to 20 kHz	41 μ V/V	
	20 kHz to 50 kHz	0.13 mV/V	
	50 kHz to 100 kHz	0.5 mV/V	



NO : SAMM 088

Page: 20 of 60

FIELD OF CALIBRATION: ELECTRICAL

SCOPE OF ACCREDITATION:

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
6. Sourcing/ Generating Instrument AC VOLTAGE (continue)	<u>1000 V @ 70 % Band Span</u> <u>(600 V to 800V)</u> 50 kHz Frequency Band Span (45 kHz to 55 kHz)	0.11 mV/V	Wavetek MTS 4950
	100 kHz Frequency Band Span (90 kHz to 110 kHz)	0.35 mV/V	
	<u>1000 V @ 100 % Band Span</u> <u>(900 V to 1100V)</u> 55 Hz Frequency Band Span (46.25 Hz to 63.75 Hz), 300 Hz Frequency Band Span (270 Hz to 440 Hz), 1 kHz Frequency Band Span (0.9 kHz to 1.1 kHz)	37 μ V/V	
	10 kHz Frequency Band Span (9 kHz to 11 kHz)	42 μ V/V	
	20 kHz Frequency Band Span (18 kHz to 22 kHz)	47 μ V/V	
	30 kHz Frequency Band Span (27 kHz to 33 kHz)	74 μ V/V	
	<u>1000 V Range(100V to 700 V)</u> 1 Hz to 40 Hz	0.4 mV/V + 0.04 V	HP 3458A Opt 002
	<u>1000V Range(600V to 1050V)</u> 10 Hz to 20 Hz 20 Hz to 40 Hz 40 Hz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz	0.2 mV/V 99 μ V/V 38 μ V/V 0.13 mV/V 0.5 mV/V	Fluke 5790A



NO : SAMM 088

Page: 21 of 60

FIELD OF CALIBRATION: ELECTRICAL

SCOPE OF ACCREDITATION:

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an -Uncertainty (\pm)*	Remarks
6. Sourcing/ Generating Instrument DC CURRENT	1 pA to +20 pA -20 pA to -1 pA	0.01 A/A + 3 nA 0.01 A/A + 3 nA	Keithley 6517/ Agilent 4339A
	20 pA to +200 pA -200 pA to -20 pA	0.01 A/A + 5 nA 0.01 A/A + 5 nA	
	200 pA to +2 nA -2 nA to -200 Pa	2 mA/A + 0.3 μ A 2 mA/A + 0.3 μ A	
	1 nA to +100 nA -100 nA to -1 nA	30 μ A/A + 0.04 nA 30 μ A/A + 0.04 nA	HP 3458A Opt 002
	100 nA to +1 μ A -1 μ A to -100 nA	20 μ A/A + 0.04 nA 20 μ A/A + 0.04 nA	
	1 μ A to +10 μ A -10 μ A to -1 μ A	20 μ A/A + 0.1 nA 20 μ A/A + 0.1 nA	
	10 μ A to +100 μ A -100 μ A to -10 μ A	20 μ A/A + 0.8 nA 20 μ A/A + 0.8 nA	
	100 μ A to +1 mA -1 mA to -100 μ A	20 μ A/A + 5 nA 20 μ A/A + 5 nA	
	110 μ A @ \pm 100 % Band Span \pm (90 μ A to 110 μ A)	21 μ A/A	Wavetek MTS 4950
	1 mA @ \pm 100 % Band Span \pm (0.9 mA to 1.1 mA) & 10 mA @ \pm 100 % Band Span \pm (9 mA to 11 mA)	11 μ A/A	
	1 mA to +10 mA -10 mA to -1 mA	20 μ A/A + 0.05 μ A 20 μ A/A + 0.05 μ A	HP 3458A Opt 002
	10 mA to +100 mA -100 mA to -10 mA	35 μ A/A + 0.5 μ A 35 μ A/A + 0.5 μ A	



NO : SAMM 088

Page: 22 of 60

FIELD OF CALIBRATION: ELECTRICAL

SCOPE OF ACCREDITATION:

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
6. Sourcing/ Generating Instrument DC CURRENT	100 mA @ ± 100 % Band Span \pm (90 mA to 110 mA)	15 μ A/A	Wavetek MTS 4950
	100 mA to +1 A -1 A to -100 mA	0.11 mA/A + 0.01 mA 0.11 mA/A + 0.01 mA	HP 3458A Opt 002
	1 A @ ± 100 % Band Span \pm (0.9 A to 1.1 A)	25 μ A/A	Wavetek MTS 4950
	+1 A to +10 A -10 A to -1 A	5.1 mA 5.1 mA	HP 3458A & Guildline 9211A.
	10 A @ ± 100 % Band Span \pm (9 A to 11 A)	55 μ A/A	Wavetek MTS 4950
	+10 A to +100 A -100 A to -10 A	0.05 A 0.05 A	HP 3458A & Guildline 9211A.
	AC CURRENT		
	5 μ A to 100 μ A 10 Hz to 20 Hz 20 Hz to 45 Hz 45 Hz to 100 Hz 100 Hz to 5 kHz	4 mA/A + 0.03 μ A 1.5 mA/A + 0.03 μ A 0.6 mA/A + 0.03 μ A 0.6 mA/A + 0.03 μ A	HP 3458A Opt 002
	100 μ A @ 100 % Band Span (90 μ A to 110 μ A) 10 Hz Frequency Band Span (9 Hz to 11 Hz) & 20 Hz Frequency Band Span (18 Hz to 22 Hz)	0.13 mA/A	Wavetek MTS 4950
	30 Hz Frequency Band Span (27 Hz to 33 Hz)	0.11 mA/A	
	40 Hz Frequency Band Span (36 Hz to 44 Hz), 55 Hz Frequency Band Span (46.25 Hz to 63.75 Hz), 300 Hz Frequency Band Span (0.27 kHz to 0.44 kHz) & 1 kHz Frequency Band Span (0.9 kHz to 1.1 kHz)	85 μ A/A	



NO : SAMM 088

Page: 23 of 60

FIELD OF CALIBRATION: ELECTRICAL

SCOPE OF ACCREDITATION:

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (±)*	Remarks	
6. Sourcing/ Generating Instrument AC CURRENT	<u>100 µA @ 100 % Band Span</u> (90 µA to 110 µA)	0.13 mA/A	Wavetek MTS 4950	
	5 kHz Frequency Band Span (4.5 kHz to 5.5 kHz)			
	10 kHz Frequency Band Span (9 kHz to 11 kHz)			
	20 kHz Frequency Band Span (18 kHz to 22 kHz)			
	30 kHz Frequency Band Span (27 kHz to 33 kHz)	0.4 mA/A	HP 3458A Opt 002	
	<u>0.05 mA to 1 mA</u>	4 mA/A + 0.2 µA 1.5 mA/A + 0.2 µA 0.6 mA/A + 0.2 µA 0.3 mA/A + 0.2 µA 0.6 mA/A + 0.2 µA 4 mA/A + 0.4 µA 5.5 mA/A + 1.5 µA		
	10 Hz to 20 Hz			
	20 Hz to 45 Hz			
	45 Hz to 100 Hz			
	100 Hz to 5 kHz			
	5 kHz to 20 kHz			
	20 kHz to 50 kHz			
	50 kHz to 100 kHz			
	<u>1 mA @ 100 % Band Span</u> (0.9 mA to 1.1 mA)	0.12 mA/A	Wavetek MTS 4950	
	10 Hz Frequency Band Span (9 Hz to 11 Hz),			
	20 Hz Frequency Band Span (18 Hz to 22 Hz) &			
	30 Hz Frequency Band Span (27 Hz to 33 Hz)			
	40 Hz Frequency Band Span (36 Hz to 44 Hz),	80 µA/A		
	55 Hz Frequency Band Span (46.25 Hz to 63.75 Hz),			
	300 Hz Frequency Band Span (0.27 kHz to 0.44 kHz) &			
	1 kHz Frequency Band Span (0.9 kHz to 1.1 kHz)			
	5 kHz Frequency Band Span (4.5 kHz to 5.5 kHz)	0.13 mA/A		



Schedule

Issue date : 12 September 2011
Valid until : 14 September 2014



NO : SAMM 088

Page: 24 of 60

FIELD OF CALIBRATION: ELECTRICAL

SCOPE OF ACCREDITATION:

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
6. Sourcing/ Generating Instrument AC CURRENT	<u>1 mA @ 100 % Band Span</u> <u>(0.9 mA to 1.1 mA)</u>		Wavetek MTS 4950
	10 kHz Frequency Band Span (9 kHz to 11 kHz)	0.26 mA/A	
	20 kHz Frequency Band Span (18 kHz to 22 kHz)	0.29 mA/A	
	30 kHz Frequency Band Span (27 kHz to 33 kHz)	0.35 mA/A	HP 3458A Opt 002
	<u>0.5 mA to 10 mA</u>		
	10 Hz to 20 Hz	4 mA/A + 2 μ A	
	20 Hz to 45 Hz	1.5 mA/A + 2 μ A	
	45 Hz to 100 Hz	0.6 mA/A + 2 μ A	
	100 Hz to 5 kHz	0.3 mA/A + 2 μ A	
	5 kHz to 20 kHz	0.6 mA/A + 2 μ A	
	20 kHz to 50 kHz	4 mA/A + 4 μ A	
	50 kHz to 100 kHz	5.5 mA/A + 15 μ A	
	<u>10 mA @ 100 % Band Span</u> <u>(9 mA to 11 mA)</u>		Wavetek MTS 4950
	10 Hz Frequency Band Span (9 Hz to 11 Hz) & 20 Hz Frequency Band Span (18 Hz to 22 Hz)	0.12 mA/A	
	30 Hz Frequency Band Span (27 Hz to 33 Hz)	96 μ A/A	
	40 Hz Frequency Band Span (36 Hz to 44 Hz), 55 Hz Frequency Band Span (46.25 Hz to 63.75 Hz), 300 Hz Frequency Band Span (0.27 kHz to 0.44 kHz) & 1 kHz Frequency Band Span (0.9 kHz to 1.1 kHz)	75 μ A/A	
	5 kHz Frequency Band Span (4.5 kHz to 5.5 kHz)	0.12 mA/A	



NO : SAMM 088

Page: 25 of 60

FIELD OF CALIBRATION: ELECTRICAL

SCOPE OF ACCREDITATION:

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
6. Sourcing/ Generating Instrument AC CURRENT	<u>10 mA @ 100 % Band Span</u> (9 mA to 11 mA)		
	10 kHz Frequency Band Span (9 kHz to 11 kHz)	0.26 mA/A	Wavetek MTS 4950
	20 kHz Frequency Band Span (18 kHz to 22 kHz)	0.29 mA/A	
	30 kHz Frequency Band Span (27 kHz to 33 kHz)	0.35 mA/A	
	<u>5 mA to 100 mA</u> 10 Hz to 20 Hz 20 Hz to 45 Hz 45 Hz to 100 Hz 100 Hz to 5 kHz 5 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz	4 mA/A + 0.02 mA 1.5 mA/A + 0.02 mA 0.6 mA/A + 0.02 mA 0.3 mA/A + 0.02 mA 0.6 mA/A + 0.02 mA 4 mA/A + 0.04 mA 5.5 mA/A + 0.15 mA	HP 3458A Opt 002
	<u>100 mA @ 100 % Band Span</u> (90 mA to 110 mA)		
	10 Hz Frequency Band Span (9 Hz to 11 Hz) & 20 Hz Frequency Band Span (18 Hz to 22 Hz)	0.12 mA/A	Wavetek MTS 4950
	30 Hz Frequency Band Span (27 Hz to 33 Hz)	96 μ A/A	
	40 Hz Frequency Band Span (36 Hz to 44 Hz), 55 Hz Frequency Band Span (46.25 Hz to 63.75 Hz), 300 Hz Frequency Band Span (0.27 kHz to 0.44 kHz) & 1 kHz Frequency Band Span (0.9 kHz to 1.1 kHz)	75 μ A/A	
	5 kHz Frequency Band Span (4.5 kHz to 5.5 kHz)	0.12 mA/A	



NO : SAMM 088

Page: 26 of 60

FIELD OF CALIBRATION: ELECTRICAL

SCOPE OF ACCREDITATION:

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
6. Sourcing/ Generating Instrument AC CURRENT	<u>100 mA @ 100 % Band Span</u> (90 mA to 110 mA)		
	10 kHz Frequency Band Span (9 kHz to 11 kHz)	0.26 mA/A	Wavetek MTS 4950
	20 kHz Frequency Band Span (18 kHz to 22 kHz)	0.29 mA/A	
	30 kHz Frequency Band Span (27 kHz to 33 kHz)	0.35 mA/A	
	<u>0.05 A to 1 A</u>		HP 3458A Opt 002
	10 Hz to 20 Hz	4 mA/A + 0.2 mA	
	20 Hz to 45 Hz	1.6 mA/A + 0.2 mA	
	45 Hz to 100 Hz	0.8 mA/A + 0.2 mA	
	100 Hz to 5 kHz	1 mA/A + 0.2 mA	
	5 kHz to 20 kHz	3 mA/A + 0.2 mA	
	20 kHz to 50 kHz	10 mA/A + 0.4 mA	
	<u>1 A @ 100 % Band Span</u> (0.9 A to 1.1 A)		Wavetek MTS 4950
	10 Hz Frequency Band Span (9 Hz to 11 Hz) & 20 Hz Frequency Band Span (18 Hz to 22 Hz)	0.17 mA/A	
	30 Hz Frequency Band Span (27 Hz to 33 Hz)	0.16 mA/A	
	40 Hz Frequency Band Span (36 Hz to 44 Hz), 55 Hz Frequency Band Span (46.25 Hz to 63.75 Hz), 300 Hz Frequency Band Span (0.27 kHz to 0.44 kHz) & 1 kHz Frequency Band Span (0.9 kHz to 1.1 kHz)	0.11 mA/A	
	5 kHz Frequency Band Span (4.5 kHz to 5.5 kHz)	0.22 mA/A	

NO : SAMM 088

Page: 27 of 60

FIELD OF CALIBRATION: ELECTRICAL

SCOPE OF ACCREDITATION:

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
6. Sourcing/ Generating Instrument AC CURRENT (continue)	10 kHz Frequency Band Span (9 kHz to 11 kHz)	0.26 mA/A	Wavetek MTS 4950
	20 kHz Frequency Band Span (18 kHz to 22 kHz)	0.29 mA/A	
	30 kHz Frequency Band Span (27 kHz to 33 kHz)	0.35 mA/A	
	<u>10 A @ 100 % Band Span</u> <u>(9 A to 11 A)</u>		
	10 Hz Frequency Band Span (9 Hz to 11 Hz),	0.24 mA/A	
	20 Hz Frequency Band Span (18 Hz to 22 Hz) &		
	30 Hz Frequency Band Span (27 Hz to 33 Hz)		
	40 Hz Frequency Band Span (36 Hz to 44 Hz)	0.22 mA/A	
	55 Hz Frequency Band Span (46.25 Hz to 63.75 Hz),	0.2 mA/A	
	300 Hz Frequency Band Span (0.27 kHz to 0.44 kHz) &		
	1 kHz Frequency Band Span (0.9 kHz to 1.1 kHz)		
	5 kHz Frequency Band Span (4.5 kHz to 5.5 kHz)	0.26 mA/A	HP 3458A & Zenith H007098.
	10 kHz Frequency Band Span (9 kHz to 11 kHz)	0.34 mA/A	
	20 kHz Frequency Band Span (18 kHz to 22 kHz)	1.2 mA/A	
	<u>1 A to 50 A</u> 50 Hz/ 60 Hz	0.5 mA/A	



NO : SAMM 088

Page: 28 of 60

*FIELD OF CALIBRATION: ELECTRICAL

*SCOPE OF ACCREDITATION:

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
6. Sourcing/ Generating Instrument RESISTANCE	0 Ω to 10 Ω 10 Ω to 100 Ω 0.1 k Ω to 1 k Ω 1 k Ω to 10 k Ω 10 k Ω to 100 k Ω 0.1 M Ω to 1 M Ω 1 M Ω to 10 M Ω 10 M Ω to 100 M Ω 100 M Ω to 1 G Ω 1 G Ω to 2 G Ω 2 G Ω to 20 G Ω 20 G Ω to 200 G Ω 0.2 T Ω to 2 T Ω 2 T Ω to 20 T Ω 20 T Ω to 200 T Ω	15 $\mu\Omega/\Omega$ + 0.05 m Ω 12 $\mu\Omega/\Omega$ + 0.5 m Ω 10 $\mu\Omega/\Omega$ + 0.5 m Ω 10 $\mu\Omega/\Omega$ + 5 m Ω 10 $\mu\Omega/\Omega$ + 0.05 Ω 15 $\mu\Omega/\Omega$ + 2 Ω 50 $\mu\Omega/\Omega$ + 0.1 k Ω 0.5 m Ω/Ω + 1 k Ω 5 m Ω/Ω + 0.01 M Ω 2.3 m Ω/Ω + 0.01 M Ω 2.3 m Ω/Ω + 0.1 M Ω 3.5 m Ω/Ω + 1 M Ω 3.5 m Ω/Ω + 0.01 G Ω 10.3 m Ω/Ω + 0.1 G Ω 11.5 m Ω/Ω + 1 G Ω	HP 3458A Opt 002 & Keithley 6517 / Agilent 4339A
	<u>10 Ω</u> 10 % Band Span (Zero to 2 Ω), 30 % Band Span (2 Ω to 4 Ω), 100 % Band Span (9 Ω to 11 Ω), 190 % Band Span (18.0 Ω to 19.5 Ω)	8 $\mu\Omega/\Omega$	Wavetek MTS 4950
	<u>100 Ω</u> 30 % Band Span (20 Ω to 40 Ω), 100 % Band Span (90 Ω to 110 Ω), 190 % Band Span (180 Ω to 195 Ω)	6 $\mu\Omega/\Omega$	
	<u>1 kΩ</u> 30 % Band Span (0.2 k Ω to 0.4 k Ω), 100 % Band Span (0.9 k Ω to 1.1 k Ω), 190 % Band Span (1.8 k Ω to 1.95 k Ω)	3 $\mu\Omega/\Omega$	

NO : SAMM 088

Page: 29 of 60

FIELD OF CALIBRATION: ELECTRICAL

SCOPE OF ACCREDITATION:

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
6. Sourcing/ Generating Instrument RESISTANCE (continue)	<u>10 kΩ</u> 30 % Band Span (2 k Ω to 4 k Ω), 100 % Band Span (9 k Ω to 11 k Ω), 190 % Band Span (18.0 k Ω to 19.5 k Ω)	3 $\mu\Omega/\Omega$	Wavetek MTS 4950
	<u>100 kΩ</u> 30 % Band Span (20 k Ω to 40 k Ω), 100 % Band Span (90 k Ω to 110 k Ω) & 190 % Band Span (180 k Ω to 195 k Ω)	6 $\mu\Omega/\Omega$	
	<u>1 MΩ</u> 30 % Band Span (0.2 M Ω to 0.4 M Ω), 100 % Band Span (0.9 M Ω to 1.1 M Ω) & 190 % Band Span (1.80 M Ω to 1.95 M Ω)	11 $\mu\Omega/\Omega$	
	<u>10 MΩ</u> 30 % Band Span (2 M Ω to 4 M Ω), 100 % Band Span (9 M Ω to 11 M Ω) & 190 % Band Span (18.0 M Ω to 19.5 M Ω)	21 $\mu\Omega/\Omega$	
	<u>100 MΩ</u> 30 % Band Span (20 M Ω to 40 M Ω), 100 % Band Span (90 M Ω to 110 M Ω)	83 $\mu\Omega/\Omega$	
CAPACITANCE	1 pF to 1 mF @ (10 Hz to 1 MHz)	0.5 mF/F	HP 4284A
INDUCTANCE	1 μ H to 10 H @ (10 Hz to 1 MHz)	0.5 mH/H	
FREQUENCY	1 Hz to 5 GHz	3.6 nHz/Hz	HP53132A



NO : SAMM 088

Page: 30 of 60

FIELD OF CALIBRATION: ELECTRICAL

SCOPE OF ACCREDITATION:

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
6. Sourcing/ Generating Instrument RPM generating instruments (Non-Contact)	10 to 29.999 rpm 30 to 59.999 rpm 60 to 99.999 rpm 100 to 299.99 rpm 300 to 599.99 rpm 600 to 999.99 rpm 1000 to 2999.9 rpm 3000 to 5999.9 rpm 6000 to 9999.9 rpm 10000 to 29999 rpm 30000 to 59999 rpm 60000 to 99999 rpm	0.014 rpm 0.036 rpm 0.07 rpm 0.14 rpm 0.37 rpm 0.7 rpm 1.4 rpm 3.6 rpm 7 rpm 14 rpm 36 rpm 70 rpm	In-House method, ESF/0303.
RPM generating instruments (Contact Measurement)	20 to 29.999 rpm 30 to 59.999 rpm 60 to 99.999 rpm 100 to 299.99 rpm 300 to 599.99 rpm 600 to 999.99 rpm 1000 to 2999.9 rpm 3000 to 5999.9 rpm 6000 to 9999.9 rpm 10000 to 12999 rpm 13000 to 25999 rpm 26000 to 29999 rpm	0.011 rpm 0.016 rpm 0.029 rpm 0.07 rpm 0.16 rpm 0.29 rpm 0.7 rpm 1.6 rpm 2.9 rpm 7 rpm 8 rpm 14 rpm	

NO : SAMM 088

Page: 31 of 60

FIELD OF CALIBRATION: ELECTRICAL

SCOPE OF ACCREDITATION:

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
7. High Voltage Tester DC VOLTAGE	0 kV to 9.999 kV 10 kV to <20 kV 20 kV to 35 kV >35 kV to 40 kV	5 mV/V + 2 V 20 mV/V 10 mV/V 20 mV/V	TDV 20 ADS Fluke 80K- 40
AC VOLTAGE	(0 kV to 9.999 kV) @ 50/60Hz (10 kV to 20 kV) @ 50/60 Hz (20 kV to 28 kV) @ 50/60 Hz	7 mV/V + 3 V 7 mV/V + 30 V 50 mV/V	

Signatories:

1. Mr. Mohd. Fahimy Bin Ahmad Ta'adin 610322-08-6029
2. Mr. Mohd. Fikri Bin Mohd Nor 660219-03-5499
3. Mr. Syahrel Bin Shari 721216-01-6029



NO : SAMM 088

Page: 32 of 60

FIELD OF CALIBRATION: ELECTRICAL

SITE CALIBRATION: CATEGORY I

SCOPE OF ACCREDITATION:

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
1. Measuring Instrument DC VOLTAGE	± 220 mV Range $\pm (0$ mV to 220 mV) ± 2.2 V Range $\pm (0$ V to 2.2 V) $+11$ V Range $\pm (0$ V to 11 V) ± 22 V Range $\pm (0$ V to 22 V) ± 220 V Range $\pm (0$ V to 220 V) ± 1100 V Range $\pm (100$ V to 1100 V)	$9 \mu\text{V/V} + 0.8 \mu\text{V}$ $8 \mu\text{V/V} + 1.2 \mu\text{V}$ $8 \mu\text{V/V} + 4 \mu\text{V}$ $8 \mu\text{V/V} + 8 \mu\text{V}$ $9 \mu\text{V/V} + 0.1 \text{ mV}$ $11 \mu\text{V/V} + 0.6 \text{ mV}$	Generating using calibrator model Fluke 5700A
AC VOLTAGE	<u>2.2 mV to 220 V</u> (See Matrix C) <u>1100 V Range</u> <u>(110 V to 1100 V)</u> 50 Hz to 1 kHz <u>800.01 V to 1050.00 V</u> 1 kHz to 3 kHz 3 kHz to 10 kHz 10 kHz to 20 kHz	(See Matrix C) $90 \mu\text{V/V} + 4 \text{ mV}$ $0.8 \text{ mV/V} + 0.13 \text{ V}$ $0.8 \text{ mV/V} + 0.21 \text{ V}$ $1.2 \text{ mV/V} + 0.32 \text{ V}$	Generating using calibrator model Wavetek 9100
DC CURRENT	$0 \mu\text{A}$ to $+220 \mu\text{A}$ $-220 \mu\text{A}$ to $0 \mu\text{A}$ 0 mA to $+2.2 \text{ mA}$ -2.2 mA to 0 mA 0 mA to $+22 \text{ mA}$ -22 mA to 0 mA 0 mA to $+220 \text{ mA}$ -220 mA to 0 mA 0 A to $+1 \text{ A}$ -1 A to 0 A $+0.32001 \text{ A}$ to $+3.2 \text{ A}$ -3.2 A to -0.32001 A $+3.2001 \text{ A}$ to $+10.5 \text{ A}$ -10.5 A to -3.2001 A $+10.5001 \text{ A}$ to $+20 \text{ A}$ -20 A to -10.5001 A $+20 \text{ A}$ to $+100 \text{ A}$ -100 A to -20 A	$60 \mu\text{A/A} + 0.01 \mu\text{A}$ $60 \mu\text{A/A} + 0.01 \mu\text{A}$ $60 \mu\text{A/A} + 0.1 \mu\text{A}$ $70 \mu\text{A/A} + 1 \mu\text{A}$ $95 \mu\text{A/A} + 30 \mu\text{A}$ $0.6 \text{ mA/A} + 0.12 \text{ mA}$ $0.55 \text{ mA/A} + 0.94 \text{ mA}$ $0.55 \text{ mA/A} + 4.5 \text{ mA}$ $0.45 \text{ mA/A} + 20 \text{ mA}$	Generating using calibrator model Fluke 5700A Generating using calibrator model Wavetek 9100 California 3213K

NO : SAMM 088

Page: 33 of 60

FIELD OF CALIBRATION: ELECTRICAL

SITE CALIBRATION: CATEGORY I

SCOPE OF ACCREDITATION:

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
1. Measuring Instrument AC CURRENT	<u>9 μA to 22 A</u> (See Matrix D)	(See Matrix D)	Generating using calibrator model Fluke 5700A
	<u>0.32001 A to 3.2 A</u> 10 Hz to 3 kHz 3 kHz to 10 kHz <u>3.2001 A to 10.5 A</u> 10 Hz to 3 kHz 3 kHz to 10 kHz <u>10.5001 A to 20 A</u> 10 Hz to 3 kHz 3 kHz to 10 kHz <u>20 A to 100 A</u> 10 Hz to 1 kHz 1 kHz to 10 kHz	1 mA/A + 0.48 mA 2.5 mA/A + 2.6 mA 2 mA/A + 3 mA 5 mA/A + 10 mA 2 mA/A + 6.9 mA 5 mA/A + 23 mA 2.5 mA/A + 0.03 A 5 mA/A + 0.05 A	Generating using calibrator model Wavetek 9100
			CA 3213K
RESISTANCE Fixed Value	10 m Ω	0.1 m Ω / Ω	Fluke 5700A & Fluke 742A Series
	100 m Ω	0.02 m Ω / Ω	
	1 Ω	5 $\mu\Omega$ / Ω	
	1.9 Ω	0.11 m Ω / Ω	
	10 Ω	5 $\mu\Omega$ / Ω	
	19 Ω	31 $\mu\Omega$ / Ω	
	100 Ω	5 $\mu\Omega$ / Ω	
	190 Ω	0.02 m Ω / Ω	
	1 k Ω	5 $\mu\Omega$ / Ω	
	1.9 k Ω	15 $\mu\Omega$ / Ω	
	10 k Ω	5 $\mu\Omega$ / Ω	
	19 k Ω	14 $\mu\Omega$ / Ω	
	100 k Ω	5 $\mu\Omega$ / Ω	
	190 k Ω	15 $\mu\Omega$ / Ω	
	1 M Ω	5 $\mu\Omega$ / Ω	
	1.9 M Ω	24 $\mu\Omega$ / Ω	
	10 M Ω	5 $\mu\Omega$ / Ω	
	19 M Ω	0.05 m Ω / Ω	
	100 M Ω	0.13 m Ω / Ω	Takeda Riken TR45
	1 G Ω	5 $\mu\Omega$ / Ω	
	10 G Ω	0.02 Ω / Ω	



NO : SAMM 088

Page: 34 of 60

FIELD OF CALIBRATION: ELECTRICAL

SITE CALIBRATION: CATEGORY I

SCOPE OF ACCREDITATION:

Matrix C

AC Voltage Measurement

Range		Frequency			
		Hz		kHz	
2.2 mV	0.22 mV to 2.2 mV	0.6 + 0.005	0.24 + 0.005	0.12 + 0.005	0.41 + 0.005
22 mV	2.2 mV to 22 mV	0.6 + 0.006	0.24 + 0.006	0.12 + 0.006	0.41 + 0.006
220 mV	22 mV to 220 mV	0.6 + 0.016	0.24 + 0.01	0.11 + 0.01	0.36 + 0.01
2.2 V	0.22 V to 2.2 V	0.6 + 0.1	0.18 + 0.03	0.085 + 0.007	0.14 + 0.02
22 V	2.2 V to 22 V	0.6 + 1	0.18 + 0.3	0.085 + 0.07	0.14 + 0.2
220 V	22 V to 220 V	0.6 + 10	0.18 + 3	0.09 + 1	0.25 + 4

Range		Frequency			
		MHz			
2.2 mV	0.22 mV to 2.2 mV	0.95 + 0.008	1.3 + 0.015	1.8 + 0.03	3.6 + 0.03
22 mV	2.2 mV to 22 mV	0.95 + 0.008	1.3 + 0.015	1.8 + 0.03	3.6 + 0.03
220 mV	22 mV to 220 mV	0.9 + 0.03	1.1 + 0.03	1.8 + 0.04	3.6 + 0.1
2.2 V	0.22 V to 2.2 V	0.28 + 0.08	0.48 + 0.15	1.2 + 0.4	2.4 + 1
22 V	2.2 V to 22 V	0.28 + 0.4	0.6 + 1.7	1.4 + 5	3 + 9
220 V	22 V to 220 V	0.6 + 10	1.6 + 110	5.4 + 110	13 + 220

The expanded uncertainties given in this table are expressed in mV/V + mV

Matrix D

AC Current Measurement

Range	Frequency (kHz)				
	0.01 to 0.02	0.02 to 0.04	0.04 to 1	1 to 5	5 to 10
9 μ A to 220 μ A	0.8 + 0.03	0.42 + 0.025	0.16 + 0.02	0.7 + 0.05	1.8 + 0.1
0.22 mA to 2.2 mA	0.8 + 0.05	0.42 + 0.04	0.16 + 0.04	0.7 + 0.5	1.8 + 1
2.2 mA to 22 mA	0.8 + 0.5	0.42 + 0.4	0.16 + 0.4	0.7 + 10	1.8 + 0.01
22 mA to 220 mA	0.8 + 5	0.42 + 4	0.18 + 4	0.7 + 50	1.8 + 100
0.22 A to 2.2 A	-	-	0.75 + 40	0.85 + 100	10 + 200

The expanded uncertainties given in this table are expressed in mA/A + μ A

NO : SAMM 088

Page: 35 of 60

FIELD OF CALIBRATION: ELECTRICAL

SITE CALIBRATION: CATEGORY I

SCOPE OF ACCREDITATION:

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
1. Measuring Instrument CAPACITANCE Fixed Value	20 Hz to 1 MHz (1, 10, 100, 1000) pF, (0.01, 0.1, 1) μ F	1 mF/F	HP 1600A series
	≤ 350 Hz 0.5 nF to 4.0 nF 4.0001 nF to 40.0 nF 40.001 nF to 400.0 nF 400.01 nF to 4.0 μ F 4.0001 μ F to 40.0 μ F 40.001 μ F to 400.0 μ F 400.01 μ F to 4.0 mF 4.0001 mF to 40.0 mF	3 mF/F + 15 pF 3 mF/F + 30 pF 3 mF/F + 0.16 nF 4 mF/F + 1.6 nF 5 mF/F + 16 nF 5 mF/F + 0.16 μ F 5 mF/F + 1.6 μ F 10 mF/F + 0.06 mF	Wavetek 9100
	350 Hz to 1.5 kHz 0.5 nF to 4.0 nF 4.0001 nF to 40.0 nF 40.001 nF to 400.0 nF 400.01 nF to 4.0 μ F 4.0001 μ F to 40.0 μ F 40.001 μ F to 400.0 μ F 400.01 μ F to 4.0 mF 4.0001 mF to 40.0 mF	6 mF/F + 0.03 nF 6 mF/F + 0.06 nF 6 mF/F + 0.32 nF 8 mF/F + 3.2 nF 10 mF/F + 32 nF 10 mF/F + 0.32 μ F 10 mF/F + 3.2 μ F 20 mF/F + 0.12 mF	
INDUCTANCE Fixed Value	100 μ H @ 1 kHz (1, 10, 100) mH @ 1 kHz 1 H @ 1 kHz	2.5 mH/H 1 mH/H 1 mH/H	ANDO RS-100 series
POWER/ ENERGY (DC)	1 kW to 20 kW 0.1 W to 1 kW	0.7 mW/W 0.22 mW/W	Fluke 5520A
POWER/ ENERGY (AC) (45 Hz to 65 Hz at PF=1)	10 kW to 20 kW 1 W to 10 kW 0.1 W to 1 W	1 mW/W 0.9 mW/W 1 mW/W	



NO : SAMM 088

Page: 36 of 60

FIELD OF CALIBRATION: ELECTRICAL

SITE CALIBRATION: CATEGORY I

SCOPE OF ACCREDITATION:

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
1. Measuring Instrument FREQUENCY	1 μ Hz to 80 MHz 100 kHz to 4320 MHz	64 nHz/Hz 32 nHz/Hz	(Agilent 33250A & R&S SMHU) & HP 53132A
Time Base Output	1 kHz to 10 MHz	3.6 nHz/Hz	HP 53132A
AMPLITUDE	10 mVp-p to 10 Vp-p (1 μ Hz to 80 MHz, Into 50 Ω)	10 mVp-p/ Vp-p + 1 mVp-p	Agilent 33250A
	+13 dBm to -140 dBm (100 kHz to 4320 MHz)	1 dBm	R&S SMHU
FLATNESS	Sine Wave Relative to 1 kHz (Auto range On)	0.4 dBm	Agilent 33250A
TIME	1 s to 24 Hrs	0.06 μ s/s + 0.021 s	Agilent 33250A & HP53132A
RPM Measuring Instruments (Non-Contact)	60 to 5999.9 rpm 6000 to 29999 rpm 30000 to 59999 rpm 60000 to 99999 rpm	0.07 rpm 7 rpm 36 rpm 70 rpm	In-House method, ESF/0303.
2. High Voltage Meter DC VOLTAGE	0.5 kV to 10 kV	5 mV/V	TDV 20 ADS & TOS 5101
AC VOLTAGE	(0.5 kV to 10 kV) @ (50/60) Hz	8 mV/V	

NO : SAMM 088

Page: 37 of 60

FIELD OF CALIBRATION: ELECTRICAL

SITE CALIBRATION: CATEGORY I

SCOPE OF ACCREDITATION:

Instrument-Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
3. Clamp Meter DC Current <u>10 – Turn Coil</u>	3.2001 A to 32 A 32.001 A to 105 A 105.001 A to 200 A	0.06 mA/A + 1.18 mA 0.55 mA/A + 9.4 mA 0.55 mA/A + 45 mA	
<u>50 – Turn Coil</u>	16.001 A to 160 A 160.01 A to 525.0 A 525.01 A to 1000.0 A	0.6 mA/A + 5.9 mA 0.55 mA/A + 47 mA 0.55 mA/A + 0.23 A	
AC Current <u>10 – Turn Coil</u>	<u>3.2001 A to 32.0 A</u> 10 Hz to 100 Hz 100 Hz to 440 Hz	2 mA/A + 5.5 mA 7.8 mA/A + 27 mA	Wavetek 9100 c/w Current Coil
	<u>32.001 A to 200.0 A</u> 10 Hz to 100 Hz 100 Hz to 440 Hz	2.1 mA/A + 90 mA 6.7 mA/A + 0.25 A	
<u>50 – Turn Coil</u>	<u>16.001 A to 160.0 A</u> 10 Hz to 100 Hz	2 mA/A + 28 mA	
	<u>160.01 A to 1000.0 A</u> 10 Hz to 100 Hz	2.1 mA/A + 0.45 A	
4. Insulation Testers	(1.0k Ω to 10.0 k Ω) @ 10.0 V, (10.0k Ω to 100.0 k Ω) @ 50.0V, (0.1M Ω to 1.0 M Ω) @ 150.0 V (1.0M Ω to 10.0M Ω) @ 300.0V, (10.0M Ω to 100.0M Ω)@500.0V (0.1G Ω to 1.0 G Ω) @ 1000.0 V (1.0G Ω to 10.0G Ω) @ 5000.0V (10.0G Ω to 100.0G Ω) @ 5000.0V (100.0G Ω to 600.0G Ω) @ 5000.0V	0.5 m Ω / Ω 1 m Ω / Ω 10 m Ω / Ω 50 m Ω / Ω 0.1 Ω / Ω 50 m Ω / Ω	Tinsley 4720



NO : SAMM 088

Page: 38 of 60

FIELD OF CALIBRATION: ELECTRICAL

SITE CALIBRATION: CATEGORY I

SCOPE OF ACCREDITATION:

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm) [*]	Remarks
5. Oscilloscope Vertical Deflection DC Signal	0V to ± 6.6 V (50 Ω Load) 0V to ± 130 V (1 M Ω Load)	2.5 mV/V + 0.04 mV 0.5 mV/V + 0.04 mV	Fluke 5500A SC600
Vertical Deflection Square Wave Signal	± 1 mVp-p to ± 6.6 Vp-p (50 Ω Load) ± 1 mVp-p to ± 130 Vp-p (1 M Ω Load)	-2.5 mVp-p/Vp-p + 0.04 mVp-p 1 mVp-p/Vp-p + 0.04 mVp-p	
Horizontal Deflection Time Markers (50 Ω Load)	2 ns/div to 20 ms/div 50 ms/div to 5 s/div	2.5 μ s/s (25+(Output \times 1000)) μ s/s	
Bandwidth Amplitude	50 kHz to 600 MHz 600 MHz to 4320 MHz	0.03 Vp-p 0.26 Vp-p	Fluke 5500A SC600 & R&H SMHU
Bandwidth Frequency	50 kHz to 600 MHz	2.5 μ Hz/Hz	Fluke 5500A SC600
	600 MHz to 4320 MHz	32 nHz/Hz	R&H SMHU &HP53132A
Risetime	≤ 300 ps	+ 0 ns / -0.1 ns	Fluke 5500A SC600

NO : SAMM 088

Page: 39 of 60

FIELD OF CALIBRATION: ELECTRICAL

SITE CALIBRATION: CATEGORY I

SCOPE OF ACCREDITATION:

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
6. Sourcing/ Generating Instrument DC VOLTAGE	± 100 mV Range $\pm (100 \mu\text{V to } 120 \text{ mV})$ ± 1 V Range $\pm (100 \text{ mV to } 1.2 \text{ V})$ ± 10 V Range $\pm (1 \text{ V to } 12 \text{ V})$ ± 100 V Range $\pm (10 \text{ V to } 120 \text{ V})$ ± 1000 V Range $\pm (100 \text{ V to } 1050 \text{ V})$	$5 \mu\text{V/V} + 0.3 \mu\text{V}$ $4 \mu\text{V/V} + 0.3 \mu\text{V}$ $4 \mu\text{V/V} + 0.5 \mu\text{V}$ $6 \mu\text{V/V} + 0.03 \text{ mV}$ $6 \mu\text{V/V} + 0.1 \text{ mV}$	HP 3458A Opt 002
AC VOLTAGE	<u>10 mV Range (1 mV to 12 mV)</u> 1 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 1 MHz 1 MHz to 4 MHz 4 MHz to 8 MHz <u>100 mV range (10 mV to 120 mV)</u> 1 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 1 MHz 1 MHz to 2 MHz 2 MHz to 4 MHz 4 MHz to 8 MHz 8 MHz to 10 MHz	$0.3 \text{ mV/V} + 3 \mu\text{V}$ $0.2 \text{ mV/V} + 1.1 \mu\text{V}$ $0.3 \text{ mV/V} + 1.1 \mu\text{V}$ $1 \text{ mV/V} + 1.1 \mu\text{V}$ $5 \text{ mV/V} + 1.1 \mu\text{V}$ $12 \text{ mV/V} + 5 \mu\text{V}$ $0.07 \text{ V/V} + 7 \mu\text{V}$ $0.2 \text{ V/V} + 8 \mu\text{V}$ $0.07 \text{ mV/V} + 4 \mu\text{V}$ $0.07 \text{ mV/V} + 2 \mu\text{V}$ $0.14 \text{ mV/V} + 2 \mu\text{V}$ $0.3 \text{ mV/V} + 2 \mu\text{V}$ $0.8 \text{ mV/V} + 2 \mu\text{V}$ $3 \text{ mV/V} + 0.01 \text{ mV}$ $0.01 \text{ V/V} + 0.01 \text{ mV}$ $15 \text{ mV/V} + 0.01 \text{ mV}$ $0.04 \text{ V/V} + 0.07 \text{ mV}$ $0.04 \text{ V/V} + 0.08 \text{ mV}$ $0.15 \text{ V/V} + 0.1 \text{ mV}$	



NO : SAMM.088

Page: 40 of 60

FIELD OF CALIBRATION: ELECTRICAL

SITE CALIBRATION: CATEGORY I

SCOPE OF ACCREDITATION:

Instrument-Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
6. Sourcing/ Generating Instrument AC VOLTAGE (Continue)	<u>1 V Range (0.1 V to 1.2 V)</u> 1 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 1 MHz 1 MHz to 2 MHz 2 MHz to 4 MHz 4 MHz to 8 MHz 8 MHz to 10 MHz <u>10 V Range (1 V to 12 V)</u> 1 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 1 MHz 1 MHz to 2 MHz 2 MHz to 4 MHz 4 MHz to 8 MHz 8 MHz to 10 MHz <u>100 V Range (10 V to 120 V)</u> 1 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz 300 kHz to 1 MHz	0.07 mV/V + 0.04 mV 0.07 mV/V + 0.02 mV 0.14 mV/V + 0.02 mV 0.3 mV/V + 0.02 mV 0.8 mV/V + 0.02 mV 3 mV/V + 0.1 mV 0.01 V/V + 0.1 mV 15 mV/V + 0.1 mV 0.04 V/V + 0.7 mV 0.04 V/V + 0.8 mV 0.15 V/V + 1 mV 0.07 mV/V + 0.4 mV 0.07 mV/V + 0.2 mV 0.14 mV/V + 0.2 mV 0.3 mV/V + 0.2 mV 0.8 mV/V + 0.2 mV 3 mV/V + 1 mV 0.01 V/V + 1 mV 15 mV/V + 1 mV 0.04 V/V + 7 mV 0.04 V/V + 8 mV 0.15 V/V + 0.01 V 0.2 mV/V + 4 mV 0.2 mV/V + 2 mV 0.2 mV/V + 2 mV 0.35 mV/V + 2 mV 1.2 mV/V + 2 mV 4 mV/V + 0.01 V 15 mV/V + 0.01 V	HP 3458A Opt 002



NO : SAMM 088

Page: 41 of 60

FIELD OF CALIBRATION: ELECTRICAL

SITE CALIBRATION: CATEGORY I

SCOPE OF ACCREDITATION:

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
6. Sourcing/-Generating Instrument AC VOLTAGE (continue)	1000 V Range (100 V to 700 V) 1 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz	0.4 mV/V + 0.04 V 0.4 mV/V + 0.02 V 0.6 mV/V + 0.02 V 1.2 mV/V + 0.02 V 3 mV/V + 0.02 V	
DC CURRENT	1 nA to +100 nA -100 nA to -1 nA 100 nA to +1 μ A -1 μ A to -100 nA 1 μ A to +10 μ A -10 μ A to -1 μ A 10 μ A to +100 μ A -100 μ A to -10 μ A 100 μ A to +1 mA -1 mA to -100 μ A 1 mA to +10 mA -10 mA to -1 mA 10 mA to +100 mA -100 mA to -10 mA 100 mA to +1 A -1 A to -100 mA +1 A to +10 A -10 A to -1 A +10 A to +100 A -100 A to -10 A +100 A to +1000 A -1000 A to -100 A	30 μ A/A + 0.04 nA 30 μ A/A + 0.04 nA 20 μ A/A + 0.04 nA 20 μ A/A + 0.04 nA 20 μ A/A + 0.1 nA 20 μ A/A + 0.1 nA 20 μ A/A + 0.8 nA 20 μ A/A + 0.8 nA 20 μ A/A + 5 nA 20 μ A/A + 5 nA 20 μ A/A + 0.05 μ A 20 μ A/A + 0.05 μ A 35 μ A/A + 0.5 μ A 35 μ A/A + 0.5 μ A 0.11 mA/A + 0.01 mA 0.11 mA/A + 0.01 mA 5.1 mA 5.1 mA 0.05 A 0.05 A 15 mA/A + 2A 15 mA/A - 2A	HP 3458A Opt 002 HP 3458A Opt 002 HP 3458A & Guildline 9211A. HP 3458A & Guildline 9211A. Kyoritsu 2003



NO : SAMM 088

Page: 42 of 60

FIELD OF CALIBRATION: ELECTRICAL

SITE CALIBRATION: CATEGORY I

SCOPE OF ACCREDITATION:

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
6. Sourcing/Generating Instrument AC CURRENT	<u>5 μA to 100 μA</u> 10 Hz to 20 Hz 20 Hz to 45 Hz 45 Hz to 100 Hz 100 Hz to 5 kHz <u>0.05 mA to 1 mA</u> 10 Hz to 20 Hz 20 Hz to 45 Hz 45 Hz to 100 Hz 100 Hz to 5 kHz 5 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz <u>0.5 mA to 10 mA</u> 10 Hz to 20 Hz 20 Hz to 45 Hz 45 Hz to 100 Hz 100 Hz to 5 kHz 5 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz	4 mA/A + 0.03 μ A 1.5 mA/A + 0.03 μ A 0.6 mA/A + 0.03 μ A 0.6 mA/A + 0.03 μ A 4 mA/A + 0.2 μ A 1.5 mA/A + 0.2 μ A 0.6 mA/A + 0.2 μ A 0.3 mA/A + 0.2 μ A 0.6 mA/A + 0.2 μ A 4 mA/A + 0.4 μ A 5.5 mA/A + 1.5 μ A 4 mA/A + 2 μ A 1.5 mA/A + 2 μ A 0.6 mA/A + 2 μ A 0.3 mA/A + 2 μ A 0.6 mA/A + 2 μ A 4 mA/A + 4 μ A 5.5 mA/A + 15 μ A	HP 3458A Opt 002



NO : SAMM 088

Page: 43 of 60

FIELD OF CALIBRATION: ELECTRICAL

SITE CALIBRATION: CATEGORY I

SCOPE OF ACCREDITATION:

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
6. Sourcing/ Generating Instrument AC CURRENT (continue)	5 mA to 100 mA 10 Hz to 20 Hz 20 Hz to 45 Hz 45 Hz to 100 Hz 100 Hz to 5 kHz 5 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 0.05 A to 1 A	4 mA/A + 0.02 mA 1.5 mA/A + 0.02 mA 0.6 mA/A + 0.02 mA 0.3 mA/A + 0.02 mA 0.6 mA/A + 0.02 mA 4 mA/A + 0.04 mA 5.5 mA/A + 0.15 mA	HP 3458A Opt 002
	10 Hz to 20 Hz 20 Hz to 45 Hz 45 Hz to 100 Hz 100 Hz to 5 kHz 5 kHz to 20 kHz 20 kHz to 50 kHz 1 A to 50 A 50 Hz/ 60 Hz	4 mA/A + 0.2 mA 1.6 mA/A + 0.2 mA 0.8 mA/A + 0.2 mA 1 mA/A + 0.2 mA 3 mA/A + 0.2 mA 10 mA/A + 0.4 mA 0.5 mA/A	HP 3458A & Zenith H007098.
	50 A to 1000 A 50 Hz/ 60 Hz 40 Hz to 1 kHz	15 mA/A + 2 A 0.03 A/A + 4 A	Kyoritsu 2003
RESISTANCE	0 Ω to 10 Ω 10 Ω to 100 Ω 0.1 k Ω to 1 k Ω 1 k Ω to 10 k Ω 10 k Ω to 100 k Ω 0.1 M Ω to 1 M Ω 1 M Ω to 10 M Ω 10 M Ω to 100 M Ω 100 M Ω to 1 G Ω	15 $\mu\Omega/\Omega$ + 0.05 m Ω 12 $\mu\Omega/\Omega$ + 0.05 m Ω 10 $\mu\Omega/\Omega$ + 0.05 m Ω 10 $\mu\Omega/\Omega$ + 5 m Ω 10 $\mu\Omega/\Omega$ + 0.05 Ω 15 $\mu\Omega/\Omega$ + 2 Ω 50 $\mu\Omega/\Omega$ + 0.1 k Ω 0.5 m Ω/Ω + 1 k Ω 5 m Ω/Ω + 0.01 M Ω	HP 3458A Opt 002



NO : SAMM 088

Page: 44 of 60

FIELD OF CALIBRATION: ELECTRICAL

SITE CALIBRATION: CATEGORY I

SCOPE OF ACCREDITATION:

Instrument-Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
6. Sourcing/ Generating Instrument CAPACITANCE	1 pF to 1 mF @ (10 Hz to 1 MHz)	0.5 mF/F	HP 4284A
INDUCTANCE	1 μ H to 10 H @ (10 Hz to 1 MHz)	0.5 mH/H	
FREQUENCY	1 Hz to 5 GHz	3.6 nHz/Hz	HP 53132A
RPM generating instruments (Non-Contact)	10 to 29.999 rpm 30 to 59.999 rpm 60 to 99.999 rpm 100 to 299.99 rpm 300 to 599.99 rpm 600 to 999.99 rpm 1000 to 2999.9 rpm 3000 to 5999.9 rpm 6000 to 9999.9 rpm 10000 to 29999 rpm 30000 to 59999 rpm 60000 to 99999 rpm	0.014 rpm 0.036 rpm 0.07 rpm 0.14 rpm 0.37 rpm 0.7 rpm 1.4 rpm 3.6 rpm 7 rpm 14 rpm 36 rpm 70 rpm	In-House method, ESF/0303.
RPM generating instruments (Contact Measurement)	20 to 29.999 rpm 30 to 59.999 rpm 60 to 99.999 rpm 100 to 299.99 rpm 300 to 599.99 rpm 600 to 999.99 rpm 1000 to 2999.9 rpm 3000 to 5999.9 rpm 6000 to 9999.9 rpm 10000 to 12999 rpm 13000 to 25999 rpm 26000 to 29999 rpm	0.011 rpm 0.016 rpm 0.029 rpm 0.07 rpm 0.16 rpm 0.29 rpm 0.7 rpm 1.6 rpm 2.9 rpm 7 rpm 8 rpm 14 rpm	



NO : SAMM 088

Page: 45 of 60

FIELD OF CALIBRATION: ELECTRICAL

SITE CALIBRATION: CATEGORY I

SCOPE OF ACCREDITATION:

Instrument-Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm) [*]	Remarks
7. High Voltage Tester DC VOLTAGE	0 kV to 9.999 kV 10 kV to <20 kV 20 kV to 35 kV >35 kV to 40 kV	5 mV/V + 2 V 20 mV/V 10 mV/V 20 mV/V	TDV 20 ADS Fluke 80K-40
AC VOLTAGE	(0 kV to 9.999 kV) @ 50/60Hz (10 kV to 20 kV) @ 50/60 Hz (20 kV to 28 kV) @ 50/60 Hz	7 mV/V + 3 V 7 mV/V + 30 V 50 mV/V	

Signatories:

1. Mr. Mohd. Fahimy Bin Ahmad Ta'adin 610322-08-6029
2. Mr. Mohd. Fikri Bin Mohd Nor 660219-03-5499
3. Mr. Syahrel Bin Shari 721216-01-6029



Schedule

Issue date : 12 September 2011
Valid until : 14 September 2014



MS ISO/IEC 17025

NO : SAMM 088

Page: 46 of 60

FIELD OF CALIBRATION: TEMPERATURE

SCOPE OF ACCREDITATION:

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)	Remarks
Liquid In-glass Thermometer (Partial Immersion)	0 °C to 200 °C	2 °C	Comparison method Using (PRT 100)

Signatory:

1. Mr. Mohd Dali Bin Silam

580306-01-5541



NO : SAMM 088

Page: 47 of 60

FIELD OF CALIBRATION: TEMPERATURE

SCOPE OF ACCREDITATION:

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
Liquid-in-glass thermometer (Total Immersion)	- 80 °C to 0 °C 0 °C to 80 °C 80 °C to 250 °C 250 °C to 500 °C	0.05 °C 0.03 °C 0.05 °C 0.10 °C	Comparison method using (PRT ,Pt100)
Platinum Resistance Thermometer	- 80 °C to 250 °C 250 °C to 600 °C 600 °C to 950 °C	0.05 °C 0.1 °C 0.5 °C	Comparison method using SPRT & PRT
Thermocouple	- 80 °C to 250 °C 250 °C to 950 °C	0.5 °C 0.6 °C	Comparison method using Platinum resistance thermometer (PRT,Pt100) and Standard Thermocouple Type R
Psychrometer Thermohygrograph Thermohygrometer	0 °C to 50 °C 35 % to 95 % RH	0.5 °C 3.0 % RH	Comparison Method Using Wet and Dry bulb base on BS 1339-1 2002
Temperature switch	- 50 °C to 10 °C 10 °C to 100 °C 100 °C to 600 °C	1 °C 2 °C 3 °C	Comparison method Using (PRT ,Pt 100 and Thermocouple Type R
Mechanical thermometer	- 80 °C to 50 °C 50 °C to 600 °C	0.1 °C 1.5 °C	Comparison method Using (PRT ,Pt 100 and Thermocouple Type R
Temperature Controlled Enclosure	- 60 °C to 200 °C 200 °C to 400 °C	0.8 °C 1.2 °C	AS 2853-1986

NO : SAMM 088

Page: 48 of 60

FIELD OF CALIBRATION: TEMPERATURE

SCOPE OF ACCREDITATION:

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
Temperature Calibrator (Dry block)-	- 30 °C to 400 °C 400 °C to 950 °C	0.1 °C 1.2 °C	EA-10 /13 (EAL Guideline for calibration of Dry block Calibrator)
Temperature Indicating Instruments			
K - type	- 200 °C to 0 °C 0 °C to 1300 °C	0.4 °C 0.4 °C	Calibration by Electrical simulation using Temperature Pressure Calibrator and Documenting Process Calibrator
J - type	- 200 °C to 0 °C 0 °C to 1100 °C	0.3 °C 0.3 °C	
T - type	- 200 °C to 0 °C 0 °C to 400 °C	0.5 °C 0.4 °C	
E - type	- 200 °C to 0 °C 0 °C to 900 °C	0.4 °C 0.3 °C	
R - type	0 °C to 1000 °C 1000 °C to 1700 °C	1.3 °C 1.1 °C	
S - type	- 50 °C to 1000 °C 1000 °C to 1700 °C	1.3 °C 1.2 °C	
Pt 100	- 200 °C to 0 °C 0 °C to 800 °C	0.2 °C 0.5 °C	

Signatories:

- | | | |
|----|---------------------------|----------------|
| 1. | Mr. Mohd Dali Bin Silam | 580306-01-5541 |
| 2. | Mr. Abdul Jalil Baharudin | 591113-10-55 |
| 3. | Mr. Zulkifli Ahmad | 701129-08-5183 |



NO : SAMM 088

Page: 49 of 60

FIELD OF CALIBRATION: TEMPERATURE

SITE CALIBRATION: CATEGORY I

SCOPE OF ACCREDITATION:

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
Humidity Chamber	20 % to 95 % RH	2.0 % of RH	AS 2853-1986 BS339-1:2002
Temperature Controlled Enclosure	- 60 °C to 200 °C 200 °C to 400 °C 400 °C to 950 °C	0.8 °C 1.2 °C 2 °C	AS 2853 -1986
Temperature Indicating Instruments			Calibration by Electrical simulation using Temperature Pressure Calibrator and Documenting Process Calibrator
K - type	- 200 °C to 0 °C 0 °C to 1300 °C	0.4 °C 0.4 °C	
J - type	- 200 °C to 0 °C 0 °C to 1100 °C	0.3 °C 0.3 °C	
T - type	- 200 °C to 0 °C 0 °C to 400 °C	0.5 °C 0.4 °C	
E - type	- 200 °C to 0 °C 0 °C to 900 °C	0.4 °C 0.3 °C	
R - type	0 °C to 1000 °C 1000 °C to 1700 °C	1.3 °C 1.1 °C	
S - type	- 50 °C to 1000 °C 1000 °C to 1700 °C	1.3 °C 1.2 °C	
Pt 100	- 200 °C to 0 °C 0 °C to 800 °C	0.2 °C 0.5 °C	

Signatories:

1. Mr. Mohd Dali Bin Silam 580306-01-5541
2. Mr. Abdul Jalil Baharudin 591113-10-55
3. Mr. Zulkifli Ahmad 701129-08-5183



Schedule

Issue date : 12 September 2011
Valid until : 14 September 2014



NO : SAMM 088

Page: 50 of 60

FIELD OF CALIBRATION: DIMENSIONAL

SCOPE OF ACCREDITATION:

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
Gauge Block	0.5 mm to 100 mm 100 mm to 250 mm	$(0.1 + 1.0 L) \mu\text{m}$ L in m	Tesa UPD
Gauge Block (Imperial units)	0.05 inch to 4 inch	$(0.004 + 1.0 L) \mu\text{inch}$ L in inch	Combination Block Angle
Bevel Protractor	0 to 360 degrees : Graduation 5 min. Graduation 1 deg.	6 minute 1 degree	Uni. Measuring M/c
Snap Gauge / Caliper Gauge	Up to 200 mm	5 μm	Uni. Measuring M/c
Micrometer	0.001 mm to 100 mm	2 μm	Gauge Block
Caliper	0.01 mm to 300 mm 0.01 mm to 600 mm 0.01 mm to 1000 mm	7 μm 9 μm 12 μm	Caliper Checker
Height Gauge	0.01 mm to 600 mm	10 μm	Caliper Checker
Caliper Checker	20 mm to 300 mm 301 mm to 600 mm	3 μm 6 μm	Long Gauge
Calibration Tester	0 mm to 25 mm	2 μm	Sylvac Probe /Gauge Block
Dial / Digital Test Indicator	0.001 mm to 100 mm	2 μm	Optimar 100
Dial / Digital Gauge	0.001 mm to 100 mm	2 μm	Optimar 100
Dial / Digital Thickness Gauge	0.001 mm to 50 mm	5 μm	Gauge Block
Depth Micro-Checker	0 mm to 300 mm	6 μm	Gauge Block / Pre. Height Gauge



NO : SAMM 088

Page: 51 of 60

FIELD OF CALIBRATION: DIMENSIONAL

SCOPE OF ACCREDITATION:

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
Feeler Gauge	0.005 mm to 2.0 mm	3 μ m	Sylvac Probe
Glass Scale	up to 200 mm 200 mm to 300 mm	3.8 μ m 5.6 μ m	Uni. Measuring M/c
Grind Gauge/Fine Gauge	0.001 mm to 0.15 mm	2 μ m	Sylvac Probe
Height master	0.001 mm to 600 mm	6 μ m	Pre. Height Gauge / Gauge Block
Ruler	0.1 mm to 2000 mm	0.1 mm	Length Comparator
Micrometer	0.1 mm to 2000 μ m	2 μ m	Gauge Block
Precision Height Gauge	0.001 mm to 600 mm	4 μ m	Gauge Block
Straight Edge	0 mm to 600 mm	1.5 μ m	Sylvac Probe / Straight edge
Tape	Up to 10 m Up to 20 m Up to 30 m Up to 50 m Up to 100 m	0.5 mm 0.7 mm 0.9 mm 1.3 mm 2.2 mm	Length Comparator
Thickness Foil	0.001 mm to 2.0 mm	3 μ m	Sylvac Probe
Setting/Standard Rod	up to 100 mm 101 mm to 200 mm 201 mm to 300 mm 301 mm to 400 mm 401 mm to 500 mm 501 mm to 600 mm	0.8 μ m 1.6 μ m 2.4 μ m 3.2 μ m 4.0 μ m 4.8 μ m	Universal Length Machine and Gauge Block
Plain Pin or Plug Gauge (Diameter Only)	0.1 mm to 50 mm	0.5 μ m	Universal Length Machine



NO : SAMM 088

Page: 52 of 60

FIELD OF CALIBRATION: DIMENSIONAL

SCOPE OF ACCREDITATION:

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
Plain Ring Gauge (Diameter Only)	Up to 50 mm 50 mm to 75 mm 75 mm to 100 mm 100 mm to 150 mm 150 mm to 200 mm	0.5 μ m 0.7 μ m 0.9 μ m 1.2 μ m 1.7 μ m	Universal Length Machine and Gauge Block

Signatories:

- | | | |
|----|---------------------------|----------------|
| 1. | Ms. Hasnah Bt Hassan | 630418-02-5624 |
| 2. | Mr. Abdul Jalil Baharudin | 591113-10-5595 |
| 3. | Mr. Mohd Hashim Effandi | 671201-01-5405 |

FIELD OF CALIBRATION: DIMENSIONAL

SITE CALIBRATION: CATEGORY I

SCOPE OF ACCREDITATION:

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
Profile Projector (X or Y axis only)	1 mm to 200 mm 201 mm to 300 mm	5 μ m 7 μ m	Glass Scale
Surface Table/Plate	2500 mm x 1600 mm	5 μ m	Levelnic / Plankator

Signatories:

- | | | |
|----|---------------------------|----------------|
| 1. | Ms. Hasnah Bt Hassan | 630418-02-5624 |
| 2. | Mr. Abdul Jalil Baharudin | 591113-10-5595 |
| 3. | Mr. Mohd Hashim Effandi | 671201-01-5405 |



NO : SAMM 088

Page: 53 of 60

FIELD OF CALIBRATION: MASS

SCOPE OF ACCREDITATION:

Instrument-Calibrated/ Measurement-Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
Standard Weights Class E2, F1, F2, M1, M2 & M3	1 mg 2 mg 5 mg 10 mg 20 mg 50 mg 100 mg 200 mg 500 mg 1 g 2 g 5 g 10 g 20 g 50 g 100 g 200 g 500 g 1 kg 2 kg 5 kg	0.002 mg 0.002 mg 0.002 mg 0.002 mg 0.003 mg 0.004 mg 0.005 mg 0.006 mg 0.008 mg 0.010 mg 0.012 mg 0.016 mg 0.020 mg 0.025 mg 0.03 mg 0.05 mg 0.10 mg 0.23 mg 0.5 mg 1.0 mg 3.0 mg	OIML R111 Comparison using ABBA Weighing scale
Standard Weights Class F1, F2, M1, M2 & M3	10 kg 20 kg 50 kg	15 mg 36 mg 75 mg	Based on Comparison Method OIML R111 Reference E2
Weight Blocks	1 mg to 100 g 100 g to 1 kg 1 kg to 10 kg 10 kg to 30 kg 30 kg to 60 kg	6 mg 2 mg 20 mg 40 mg 90 mg	
Spring /Hanging Scale	1 g to 5 kg 5 kg to 20 kg 20 kg to 50 kg 50 kg to 100 kg	16 g 31 g 62 g 124 g	

Signatories:

1. Mr. Abdul Jalil Baharudin 591113-10-5595
2. Mr. Mohammad Najib Kamaruddin 740507-08-5239



NO : SAMM 088

Page: 54 of 60

FIELD OF CALIBRATION: MASS

SITE CALIBRATION: CATEGORY I

SCOPE OF ACCREDITATION:

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
Analytical Balance	0.01 mg to 50 g 0.01 mg to 100 g 0.1 mg to 200 g	0.05 mg 0.08 mg 0.15 mg	Reference E2
Balance / Scale (Triple Beam, Single Beam Spring, Top Pan, Mechanical & Electronic)	0.001 g to 1000 g 0.01 g to 2000 g 0.01 g to 5000 g 0.1 g to 10000 g 0.1 g to 20 kg 0.001 kg to 50 kg 0.001 kg to 100 kg 0.001 kg to 200 kg 0.01 kg to 500 kg	0.001 g 0.01 g 0.01 g 0.06 g 0.1 g 1 g 7 g 10 g 20 g	Reference F1 Reference M2
Electrical and Mechanical Platform Balance	0.01 kg to 1000 kg 0.1 kg to 2000 kg	60 g 130 g	Reference M2

Signatories:

- | | | |
|----|-------------------------------|----------------|
| 1. | Mr. Abdul Jalil Baharudin | 591113-10-5595 |
| 2. | Mr. Mohammad Najib Kamaruddin | 740507-08-5239 |



NO : SAMM 088

Page: 56 of 60

FIELD OF CALIBRATION: FORCE

SITE CALIBRATION: CATEGORY I

SCOPE OF ACCREDITATION:

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
Universal Testing Machine / Compression mode (tension mode up to 5000kgf)	0 kgf to 100 kgf 101 kgf to 500 kgf 501 kgf to 1000 kgf 1001 kgf to 5000 kgf 5001 kgf to 10000 kgf 10001 kgf to 60000 kgf 60001 kgf to 200000 kgf	0.01 kgf 0.1 kgf 0.6 kgf 2 kgf 20 kgf 60 kgf 400 kgf	BS EN ISO 7500- 1:1999(E) LC 100 kgf LC 500 kgf LC 1000 kgf LC 5000 kgf LC 10000 kgf LC 60000 kgf LC 200000 kgf
Charpy / Izod Impact Tester	1 Joule to 300 Joules	1 % of reading	BS 131-7 : 1998

Note : LC (Load Cell)

Signatories:

- | | | |
|----|---------------------------|----------------|
| 1. | Mr. Abdul Jalil Baharudin | 5911 |
| | | 13-10-5595 |
| 2. | Mr. Noor Azam Ismail | 730629-14-5579 |



NO : SAMM 088

Page: 57 of 60

FIELD OF CALIBRATION: PRESSURE

SCOPE OF ACCREDITATION:

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
Pressure measuring device (Gas medium)	1 psi to 100 psi 101 psi to 1600 psi	0.01 % of pressure 0.01 % of pressure	Comparison with Dead Weight Tester based on BS EN 837-1, 837-2, 837-3 ADWT ADWT
Pressure measuring device (Oil medium)	10 psi to 800 psi 801 psi to 16000 psi 16 0001 psi to 40000 psi	0.04 % of pressure 0.01 % of pressure 0.02 % of pressure	DWT DWT Based on BS EN 837-1 BS EN 837-2 BS EN 837-3
Vacuum gauge	Ambient to 67 cm Hg	0.1 % of pressure	Mercury Manometer
Low Pressure Measuring Devices Mercury Manometer	0 psi to 1500 mm Hg	2 mm Hg	Reference Pressure Balance
Water Manometer	0 psi to 1500 mm H ₂ O	3 mm H ₂ O	
Deadweight Tester	10 psi to 40 000 psi	0.02% of pressure	Cross float method OIML R110 1994(E) NCSL RISP-4 2000 EAL G26/EA-4/17. JULY 1997 Pressure Balance

Note : ADWT (Air Dead Weight Tester)
DWT (Dead Weight Tester)

Signatories:

1. Mr. Abdul Jalil Baharudin 591113-10-5595
2. Mr. Noor Azam Ismail 730629-14-5579



Schedule

Issue date : 12 September 2011
Valid until : 14 September 2014



MS ISO/IEC 17025

NO : SAMM 088

Page: 58 of 60

FIELD OF CALIBRATION: PRESSURE

SITE CALIBRATION: CATEGORY I

SCOPE OF ACCREDITATION:

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)	Remarks
Pressure measuring device (Oil medium)	0 psi to 500 psi 0 psi to 1000 psi 0 psi to 3000 psi 0 psi to 4500 psi 0 psi to 7500 psi 0 psi to 10 000 psi 0 psi to 12 000 psi 0 psi to 16 000 psi	0.7 % of pressure 0.7 % of pressure 0.6 % of pressure 0.4 % of pressure 0.4 % of pressure 0.7 % of pressure 0.6 % of pressure 0.6 % of pressure	BS EN 837-1 BS EN 837-2 BS EN 837-3 Pressure Balance
Vacuum Measuring Devices	Ambient to -70 cm Hg	1.0 % of pressure	Pressure Calibrator
Pressure Gauge	0 bar to 20 bar	0.5 % of pressure	Pressure Calibrator

Signatories:

1. Mr. Abdul Jalil Baharudin 591113-10-5595
2. Mr. Noor Azam Ismail 730629-14-5579



NO : SAMM 088

Page: 59 of 60

FIELD OF CALIBRATION: VOLUMETRIC

SCOPE OF ACCREDITATION:

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
Bulb Pipette	1 ml to 5 ml 5 ml to 100 ml	0.008 ml 0.03 ml	Cal .Procedure No TCP 0002 REV2.0 Ref: ISO 4748
Graduated Pipette	1 ml to 5 ml 5 ml to 100 ml	0.008 ml 0.03 ml	Cal .Procedure No TCP 0002 REV2.0 Ref: ISO 4748
Measuring Cylinder (5ml to 2000ml)	5 ml to 100 ml 100 ml to 500 ml 500 ml to 1000 ml 1000 ml to 2000 ml	0.05 ml 0.5 ml 2.0 ml 3.0 ml	Cal .Procedure No TCP 0001 REV2.0 Ref: ISO 4748
Burette	1 ml to 50 ml	0.05 ml	Cal .Procedure No TCP 0003 REV2.0 Ref: ISO 4748
Beaker & Flasks (5ml to 2000ml)	5 ml to 100 ml 100 ml to 1000 ml 1000 ml to 2000 ml	1.0 ml 3.0 ml 5.0 ml	Cal .Procedure No TCP 0004 REV2.0 Ref: ISO 4748

Signatory:

1. Mr. Mohd Dali Bin Silam

580306-01-5541



NO : SAMM 088

Page: 60 of 60

FIELD OF CALIBRATION : HYDROMETER

SCOPE OF ACCREDITATION :

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
Density Hydrometer	0.600 g/ml to 0.650 g/ml 0.650 g/ml to 0.700 g/ml 0.700 g/ml to 0.750 g/ml 0.750 g/ml to 0.800 g/ml 0.800 g/ml to 0.850 g/ml 0.850 g/ml to 0.900 g/ml 0.900 g/ml to 0.950 g/ml 0.950 g/ml to 1.000 g/ml 1.000 g/ml to 1.050 g/ml 1.050 g/ml to 1.100 g/ml 1.100 g/ml to 1.150 g/ml 1.150 g/ml to 1.200 g/ml 1.200 g/ml to 1.250 g/ml 1.250 g/ml to 1.300 g/ml 1.300 g/ml to 1.350 g/ml 1.350 g/ml to 1.400 g/ml 1.400 g/ml to 1.450 g/ml 1.450 g/ml to 1.500 g/ml	0.0006 g/ml (graduation 0.0005g/ml)	Comparison Method BS 718:1991

Signatory:

1. Mr. Mohd Dali Bin Silam



NO : SAMM 088

Page: 55 of 60

FIELDS OF CALIBRATION: FORCE & TORQUE

SCOPE OF ACCREDITATION:

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty (\pm)*	Remarks
Force Measurement			
Proving Rings/Load Cells	0 kgf to 100 kgf 101 kgf to 500 kgf 501 kgf to 1000 kgf 1001 kgf to 5000 kgf 5001 kgf to 10000 kgf 10001 kgf to 30000 kgf	0.01 kgf 0.1 kgf 0.6 kgf 2 kgf 20 kgf 60 kgf	ISO 376 : 1999 LC 100 kgf LC 500 kgf LC 1000 kgf LC 5000 kgf LC 10000 kgf LC 45000 kgf
Push-pull/ Force Gauge	1 gf to 5 kgf 5 kgf to 20 kgf 20 kgf to 50 kgf 50 kgf to 100 kgf	16 gf 31 gf 62 gf 124 gf	ASTM E74-95 Reference F1 & F2
Gramme/ Dial Tension Gauge	1 gf to 50 gf 50 gf to 500 gf 500 gf to 2000 gf	1 gf 6 gf 30 gf	ASTM E74-95 Reference F1 & F2
Torque Measurement			
Torque wrench	0 Nm to 1 Nm 1 Nm to 50 Nm 51 Nm to 100 Nm 101 Nm to 500 Nm	0.07 Nm 0.04 Nm 0.3 Nm 1 Nm	Comparison with torque transducers with reference to BS EN ISO 6789 : 2003. Transducer 2 Nm Transducer 50 Nm Transducer 100 Nm Transducer 500 Nm
Torque meter and analyser	0 Nm to 1 Nm 1 Nm to 50 Nm 51 Nm to 100 Nm 101 Nm to 500 Nm	0.005 Nm 0.045 Nm 0.3 Nm 0.4 Nm	BS 7882 : 1997 Reference F2 & M2

Note : LC (Load Cell)

Signatories:

1. Mr. Abdul Jalil Baharudin 591113-10-5595
2. Mr. Noor Azam Ismail 730629-14-5579

